



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB128956

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	12/27/2023	16:10
2	CAL2	1	Water	NA	NA	20.4	7.01	12/27/2023	16:11
3	CAL3	1	Water	NA	NA	20.4	10.02	12/27/2023	16:15
4	ICV	1	Water	NA	NA	20.5	7.01	12/27/2023	16:20
5	CCV1	1	Water	NA	NA	20.5	2.02	12/27/2023	16:25
6	O6031-02	1	Solid	20.02	20	21.5	9.75	12/27/2023	16:35
7	O6031-02DUP	1	Solid	20.03	20	21.6	9.76	12/27/2023	16:36
8	O6031-04	1	Solid	20.01	20	21.6	6.37	12/27/2023	16:39
9	CCV2	1	Water	NA	NA	20.3	12.02	12/27/2023	16:42

VB 128956

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY O6031 WorkList ID : 176801 Department : Wet-Chemistry Date : 12-27-2023 14:48:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6031-02	MR-321-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/26/2023	9045D
O6031-04	MR-321-10-12	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/26/2023	9045D

Date/Time 12/27/23 15:00
Raw Sample Received by: SDC/CC
Raw Sample Relinquished by: SDC/CC

Date/Time 12/27/23 18:00
Raw Sample Received by: SDC/CC
Raw Sample Relinquished by: SDC/CC