

LB128959


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45239



A= 0.0000e+000

B= 2.2191e-004

C= -4.8930e-002

Rho= 0.9999786

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	-0.038	-0.038	47	0.000	47	0				14
0.2	0.200	0.228	0.028	1247	0.0 %	1247					1
2.5	2.500	2.522	0.022	11587	0.0 %	11587					0
5.0	5.000	4.981	-0.019	22667	0.0 %	22667					0
7.5	7.500	7.517	0.017	34094	0.0 %	34094					0
10.0	10.000	9.990	-0.010	45239	0.0 %	45239					0

LB128959 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Dec 2023 09:23:10

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7470A	PPB	28 Dec 2023 09:34:40	S	Std
0.2 - 1	50.2	1247	-	0.2000	7470A	PPB	28 Dec 2023 09:36:58	S	Std
2.5 - 1	52.5	11587	-	2.5000	7470A	PPB	28 Dec 2023 09:39:16	S	Std
5.0 - 1	55	22667	-	5.0000	7470A	PPB	28 Dec 2023 09:41:34	S	Std
7.5 - 1	57.5	34094	-	7.5000	7470A	PPB	28 Dec 2023 09:46:28	S	Std
10.0 - 1	59	45239	-	10.0000	7470A	PPB	28 Dec 2023 09:55:47	S	Std
ICV43 - 1	ICV43	19670	4.3161	-	7470A	PPB	28 Dec 2023 10:03:46	U	SMPL
ICB43 - 1	ICB43	146	-0.0165	-	7470A	PPB	28 Dec 2023 10:08:36	U	SMPL
CCV78 - 1	CCV78	23116	5.0808	-	7470A	PPB	28 Dec 2023 10:10:51	U	SMPL
CCB78 - 1	CCB78	-111	-0.0736	-	7470A	PPB	28 Dec 2023 10:13:06	U	SMPL
CRA - 1	CRA	1208	0.2191	-	7470A	PPB	28 Dec 2023 10:15:25	U	SMPL
HighStd - 1	HighStd	45801	10.1149	-	7470A	PPB	28 Dec 2023 10:17:41	U	SMPL
ChkStd - 1	ChkStd	32078	7.0696	-	7470A	PPB	28 Dec 2023 10:19:56	U	SMPL
PB158121BS - 1	LCSW	19170	4.2051	-	7470A	PPB	28 Dec 2023 10:24:36	U	SMPL
O5985-17 - 1	MCOLK8	4811	1.0187	-	7470A	PPB	28 Dec 2023 10:26:53	U	SMPL
O5985-18 - 1	MCOLS7	76	-0.0321	-	7470A	PPB	28 Dec 2023 10:29:12	U	SMPL
O5985-19 - 1	MCOLS8	189	-0.0070	-	7470A	PPB	28 Dec 2023 10:31:30	U	SMPL
O5985-20 - 1	MCOLS9	164	-0.0125	-	7470A	PPB	28 Dec 2023 10:33:47	U	SMPL
O5985-21 - 1	MCOLT0	164	-0.0125	-	7470A	PPB	28 Dec 2023 10:36:03	U	SMPL
CCV79 - 1	CCV79	23586	5.1851	-	7470A	PPB	28 Dec 2023 10:38:19	U	SMPL
CCB79 - 1	CCB79	-151	-0.0824	-	7470A	PPB	28 Dec 2023 10:40:34	U	SMPL
O5985-21DUP - 1	MCOLT0DUP	154	-0.0148	-	7470A	PPB	28 Dec 2023 10:42:54	U	SMPL
O5985-22 - 1	O5985-21MS	14877	3.2525	-	7470A	PPB	28 Dec 2023 10:45:09	U	SMPL
O5985-23 - 1	O5985-21MSD	16035	3.5094	-	7470A	PPB	28 Dec 2023 10:47:26	U	SMPL
O5985-24 - 1	MCOLT1	150	-0.0156	-	7470A	PPB	28 Dec 2023 10:49:43	U	SMPL
O5985-25 - 1	MCOM99	1746	0.3390	-	7470A	PPB	28 Dec 2023 10:52:01	U	SMPL
O5985-26 - 1	MCOMA0	169	-0.0114	-	7470A	PPB	28 Dec 2023 10:54:18	U	SMPL
O5985-27 - 1	MCOMA1	207	-0.0030	-	7470A	PPB	28 Dec 2023 10:56:35	U	SMPL
O5985-28 - 1	MCOMA2	86	-0.0298	-	7470A	PPB	28 Dec 2023 10:58:52	U	SMPL
O5985-29 - 1	MCOMA3	94	-0.0281	-	7470A	PPB	28 Dec 2023 11:01:09	U	SMPL
O5985-30 - 1	MCOMA4	138	-0.0183	-	7470A	PPB	28 Dec 2023 11:03:25	U	SMPL
CCV80 - 1	CCV80	23511	5.1685	-	7470A	PPB	28 Dec 2023 11:05:41	U	SMPL
CCB80 - 1	CCB80	-149	-0.0820	-	7470A	PPB	28 Dec 2023 11:07:56	U	SMPL
O5985-31 - 1	MCOMA5	113	-0.0239	-	7470A	PPB	28 Dec 2023 11:10:15	U	SMPL
O5985-32 - 1	MCOMA6	30	-0.0423	-	7470A	PPB	28 Dec 2023 11:12:32	U	SMPL
PB158122BL - 1	PBW	62	-0.0352	-	7470A	PPB	28 Dec 2023 11:14:48	U	SMPL
PB158122BS - 1	LCSW	16189	3.5436	-	7470A	PPB	28 Dec 2023 11:17:04	U	SMPL
O4699-07 - 1	LOD-MDL-WATER-01-QT4-2023	378	0.0350	-	7470A	PPB	28 Dec 2023 11:24:36	U	SMPL
O4699-07RE - 1	LOD-MDL-WATER-01-QT4-2023RE	923	0.1559	-	7470A	PPB	28 Dec 2023 11:26:52	U	SMPL
O4699-09 - 1	MDL-WATER-03-QT4-2023	1271	0.2331	-	7470A	PPB	28 Dec 2023 11:29:09	U	SMPL
O5992-01 - 1	FM406	1689	0.3259	-	7470A	PPB	28 Dec 2023 11:34:21	U	SMPL
O6014-01 - 1	AUD-1304	826	0.1344	-	7470A	PPB	28 Dec 2023 11:36:38	U	SMPL
O6024-01 - 1	A2421-FRAC-TANK	414	0.0428	-	7470A	PPB	28 Dec 2023 11:38:54	U	SMPL
CCV81 - 1	CCV81	23527	5.1720	-	7470A	PPB	28 Dec 2023 11:41:09	U	SMPL
CCB81 - 1	CCB81	-47	-0.0594	-	7470A	PPB	28 Dec 2023 11:43:25	U	SMPL
O6024-01DUP - 1	A2421-FRAC-TANKDUP	449	0.0507	-	7470A	PPB	28 Dec 2023 11:45:43	U	SMPL
O6024-01MS - 1	A2421-FRAC-TANKMS	15154	3.3139	-	7470A	PPB	28 Dec 2023 11:47:58	U	SMPL
O6024-01MSD - 1	A2421-FRAC-TANKMSD	14980	3.2753	-	7470A	PPB	28 Dec 2023 11:50:14	U	SMPL
O6024-03 - 1	TOTE-COMP	-18	-0.0529	-	7470A	PPB	28 Dec 2023 11:52:32	U	SMPL
PB158121BL - 1	PBW	81	-0.0310	-	7470A	PPB	28 Dec 2023 11:54:51	U	SMPL
O5985-21LX5 - 1		28	-0.0427	-	7470A	PPB	28 Dec 2023 11:57:08	U	SMPL
O5985-21A - 1		18675	4.0953	-	7470A	PPB	28 Dec 2023 11:59:24	U	SMPL
O6024-01LX5 - 1		-94	-0.0698	-	7470A	PPB	28 Dec 2023 12:01:41	U	SMPL
O6024-01A - 1		18918	4.1492	-	7470A	PPB	28 Dec 2023 12:04:00	U	SMPL
CCV82 - 1	CCV82	24043	5.2865	-	7470A	PPB	28 Dec 2023 12:06:16	U	SMPL
CCB82 - 1	CCB82	-147	-0.0816	-	7470A	PPB	28 Dec 2023 12:08:34	U	SMPL