



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB128967

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/28/2023	13:35
2	CAL2	1	Water	NA	NA	20.3	7.00	12/28/2023	13:36
3	CAL3	1	Water	NA	NA	20.3	10.02	12/28/2023	13:37
4	ICV	1	Water	NA	NA	20.5	7.01	12/28/2023	13:40
5	CCV1	1	Water	NA	NA	20.5	2.01	12/28/2023	13:45
6	O6038-07	1	Water	NA	NA	20.2	7.75	12/28/2023	13:50
7	O6038-07DUP	1	Water	NA	NA	20.3	7.76	12/28/2023	13:51
8	CCV2	1	Water	NA	NA	20.3	12.02	12/28/2023	14:00

VB 128967

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 0603

WorkList ID : 176833

Department : Wet-Chemistry

Date : 12-28-2023 13:24:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6038-07	EQ-TANK	Water	pH	Cool 4 deg C	PSEG03	O31	12/28/2023	904DC

Date/Time 12/28/23 13:30

Raw Sample Received by: JO CDC

Raw Sample Relinquished by: JTCsm

Date/Time 12/28/23

Raw Sample Received by: JTCsm

Raw Sample Relinquished by: JO CDC

171.00