



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 12/29/2023 16:00 TEMP1 OUT: 103 °C 01/02/2024 07:30
TEMP2 IN: 103 °C 01/02/2024 08:00 TEMP2 OUT: 104 °C 01/02/2024 09:30
TEMP3 IN: 540 °C 01/02/2024 10:00 TEMP3 OUT: 540 °C 01/02/2024 11:30
TEMP4 IN: 560 °C 01/02/2024 12:00 TEMP4 OUT: 560 °C 01/02/2024 13:30

Run Number: LB128983

SUPERVISOR: Iwona

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB128983BL	74.5963	74.5963	100	74.5963	74.5963	74.5963	74.5963	0.0000	0
2	O6038-07	85.0730	85.0730	100	85.1581	85.1581	85.1356	85.1356	0.0225	225
3	O6038-07DUP	97.3170	97.3170	100	97.4023	97.4023	97.3791	97.3791	0.0232	232

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

126983

WorkList Name : tvs i6038 water

WorkList ID : 176862

Department : Wet-Chemistry

Date : 12-29-2023 10:19:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6038-07	EQ-TANK	Water	TVS	Cool 4 deg C	PSEG03	O31	12/28/2023	SM2540 G

Date/Time 12/29/23 12:30
 Raw Sample Received by: JTC (WC1)
 Raw Sample Relinquished by: JTC (WC1)

Date/Time 12/29/23 16:10
 Raw Sample Received by: JTC (WC1)
 Raw Sample Relinquished by: JTC (WC1)