



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB128989

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/29/2023	14:30
2	CAL2	1	Water	NA	NA	20.3	7.00	12/29/2023	14:31
3	CAL3	1	Water	NA	NA	20.3	10.02	12/29/2023	14:35
4	ICV	1	Water	NA	NA	20.5	7.02	12/29/2023	14:40
5	CCV1	1	Water	NA	NA	20.4	2.02	12/29/2023	15:15
6	O6041-01	1	Solid	20.02	20	23.1	7.03	12/29/2023	15:25
7	O6041-01DUP	1	Solid	20.03	20	23.2	7.04	12/29/2023	15:26
8	CCV2	1	Water	NA	NA	20.3	12.02	12/29/2023	15:30

WORKLIST(Hardcopy Internal Chain)

VB28989

WorkList Name : ph o6041 WorkList ID : 176867 Department : Wet-Chemistry Date : 12-29-2023 14:17:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6041-01	F- BU-02-12282023	Solid	pH	Cool 4 deg C	PSEG05	O31	12/29/2023	9045D

Date/Time 12/29/23 14:20
 Raw Sample Received by: JH 1001
 Raw Sample Relinquished by: RJC EST-1001

Date/Time 12/29/23 17:30
 Raw Sample Received by: RJC EST-1001
 Raw Sample Relinquished by: JH CDC