



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB128990
BalanceID: WC SC-4

Analyst By : JIGNESH
Supervisor Review By : Iwona
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3033
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W3005
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/29/2023	14:30
2	CAL2	1	Water	NA	NA	20.3	7.00	12/29/2023	14:31
3	CAL3	1	Water	NA	NA	20.3	10.02	12/29/2023	14:35
4	ICV	1	Water	NA	NA	20.5	7.02	12/29/2023	14:40
5	CCV1	1	Water	NA	NA	20.3	2.01	12/29/2023	14:45
6	O6045-02	1	Solid	20.02	20	23.9	6.75	12/29/2023	14:50
7	O6045-02DUP	1	Solid	20.03	20	23.8	6.76	12/29/2023	14:51
8	O6045-04	1	Solid	20.02	20	24.2	6.37	12/29/2023	15:00
9	CCV2	1	Water	NA	NA	20.3	12.02	12/29/2023	15:05

~~12/28/23~~
12/28/23

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 06045 WorkList ID : 176868 Department : Wet-Chemistry Date : 12-29-2023 14:18:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6045-02	MR-300-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/28/2023	9045D
O6045-04	MR-300-2.0-2.5	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	12/28/2023	9045D

Date/Time 12/29/23 14:20
 Raw Sample Received by: RS (Est-106)
 Raw Sample Relinquished by: RS (Est-106)

Date/Time 12/29/23 17:30
 Raw Sample Received by: RS (Est-106)
 Raw Sample Relinquished by: RS (Est-106)