


LB128992

SFAM01.1

INSTRUMENT ID: CV1

μ Abs.:

51669



Linear

A= 0.0000e+000

B= 1.9527e-004 slope

C= -9.4793e-002 $y = mx + c$ intercept

Rho= 0.9997294

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.019	0.019	581	0.000	581					
0.05	0.050					442	0				8
0.20	0.200	0.215	0.015	1586	0.0 %	1586					2
2.50	2.500	2.544	0.044	13515	0.0 %	13515					-4
5.00	5.000	4.824	-0.176	25191	0.0 %	25191					1
7.50	7.500	7.604	0.104	39425	0.0 %	39425					0
10.0	10.000	9.994	-0.006	51669	0.0 %	51669					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	581	-		0 SFAM01.1	PPB	12/29/2023 16:31	Std
	0.2 S01	1586	-		0.2 SFAM01.1	PPB	12/29/2023 16:33	Std
	2.5 S02	13515	-		2.5 SFAM01.1	PPB	12/29/2023 16:36	Std
	5 S03	25191	-		5 SFAM01.1	PPB	12/29/2023 16:38	Std
	7.5 S04	39425	-		7.5 SFAM01.1	PPB	12/29/2023 16:40	Std
	10 S05	51669	-		10 SFAM01.1	PPB	12/29/2023 16:43	Std
ICV046	ICV046	20227	3.8549 -		SFAM01.1	PPB	12/29/2023 16:46	SMPL
ICB046	ICB046	478	-0.0015 -		SFAM01.1	PPB	12/29/2023 16:48	SMPL
CCV091	CCV091	25528	4.89 -		SFAM01.1	PPB	12/29/2023 16:50	SMPL
CCB091	CCB091	448	-0.0073 -		SFAM01.1	PPB	12/29/2023 16:52	SMPL
PB158170BL	PBS170	681	0.0382 -		SFAM01.1	PPB	12/29/2023 16:55	SMPL
O5814-01	EOZ06	2422	0.3781 -		SFAM01.1	PPB	12/29/2023 16:57	SMPL
O5814-02	EOZ07	2348	0.3637 -		SFAM01.1	PPB	12/29/2023 16:59	SMPL
O5814-03	EOZ08	1514	0.2008 -		SFAM01.1	PPB	12/29/2023 17:01	SMPL
O5814-04	EOZ09	633	0.0288 -		SFAM01.1	PPB	12/29/2023 17:04	SMPL
O5814-05	EOZ10	474	-0.0022 -		SFAM01.1	PPB	12/29/2023 17:06	SMPL
O5814-06	EOZ34	1144	0.1286 -		SFAM01.1	PPB	12/29/2023 17:08	SMPL
O5814-07	EOZ35	837	0.0686 -		SFAM01.1	PPB	12/29/2023 17:11	SMPL
O5814-08	EOZ31	2117	0.3186 -		SFAM01.1	PPB	12/29/2023 17:13	SMPL
O5814-09	EOZ32	3186	0.5273 -		SFAM01.1	PPB	12/29/2023 17:15	SMPL
O5814-10	EOZG1	1660	0.2294 -		SFAM01.1	PPB	12/29/2023 17:17	SMPL
O5814-11	EOZG2	2308	0.3559 -		SFAM01.1	PPB	12/29/2023 17:20	SMPL
O5814-12	EOZG3	719	0.0456 -		SFAM01.1	PPB	12/29/2023 17:22	SMPL
O5814-13	EOZG4	635	0.0292 -		SFAM01.1	PPB	12/29/2023 17:24	SMPL
O5814-14	EOZG5	496	0.0021 -		SFAM01.1	PPB	12/29/2023 17:26	SMPL
O5814-15	EOZ33	2507	0.3947 -		SFAM01.1	PPB	12/29/2023 17:29	SMPL
O5814-16	EOZK6	1800	0.2567 -		SFAM01.1	PPB	12/29/2023 17:31	SMPL
O5814-17	EOZK7	1603	0.2182 -		SFAM01.1	PPB	12/29/2023 17:33	SMPL
O5814-18	EOZK8	643	0.0308 -		SFAM01.1	PPB	12/29/2023 17:35	SMPL
O5814-19	EOZK9	644	0.031 -		SFAM01.1	PPB	12/29/2023 17:38	SMPL
CCV092	CCV092	25807	4.9445 -		SFAM01.1	PPB	12/29/2023 17:40	SMPL
CCB092	CCB092	569	0.0163 -		SFAM01.1	PPB	12/29/2023 17:42	SMPL
O5814-20	EOZL0	487	0.0003 -		SFAM01.1	PPB	12/29/2023 17:45	SMPL
O5814-21	EOZL0D	312	-0.0339 -		SFAM01.1	PPB	12/29/2023 17:47	SMPL
O5814-22	EOZL0S	13069	2.4571 -		SFAM01.1	PPB	12/29/2023 17:49	SMPL
PB158171BL	PBS171	635	0.0292 -		SFAM01.1	PPB	12/29/2023 17:51	SMPL
O5815-01	EOZD9	50578	9.7814 -		SFAM01.1	PPB	12/29/2023 17:54	SMPL
O5815-02	EOZE0	3319	0.5533 -		SFAM01.1	PPB	12/29/2023 17:56	SMPL
O5815-03	EOZE1	2428	0.3793 -		SFAM01.1	PPB	12/29/2023 17:58	SMPL
O5815-04	EOZE2	2138	0.3227 -		SFAM01.1	PPB	12/29/2023 18:00	SMPL
O5815-05	EOZE3	1974	0.2907 -		SFAM01.1	PPB	12/29/2023 18:03	SMPL
O5815-06	EOZE4	1333	0.1655 -		SFAM01.1	PPB	12/29/2023 18:05	SMPL
O5815-07	EOZE5	7016	1.2752 -		SFAM01.1	PPB	12/29/2023 18:07	SMPL
O5815-08	EOZ21	9780	1.8149 -		SFAM01.1	PPB	12/29/2023 18:10	SMPL
O5815-09	EOZ22	9417	1.744 -		SFAM01.1	PPB	12/29/2023 18:12	SMPL
O5815-10	EOZ23	7930	1.4537 -		SFAM01.1	PPB	12/29/2023 18:14	SMPL

O5815-11	E0Z24	5550	0.9889 -	SFAM01.1	PPB	12/29/2023 18:16	SMPL
O5815-12	E0Z25	8919	1.6468 -	SFAM01.1	PPB	12/29/2023 18:19	SMPL
O5815-13	E0ZD6	8529	1.5706 -	SFAM01.1	PPB	12/29/2023 18:21	SMPL
O5815-14	E0ZD7	5426	0.9647 -	SFAM01.1	PPB	12/29/2023 18:23	SMPL
O5815-15	E0ZD8	12710	2.387 -	SFAM01.1	PPB	12/29/2023 18:26	SMPL
O5815-16	E0ZF6	5402	0.96 -	SFAM01.1	PPB	12/29/2023 18:28	SMPL
CCV093	CCV093	26266	5.0341 -	SFAM01.1	PPB	12/29/2023 18:30	SMPL
CCB093	CCB093	574	0.0173 -	SFAM01.1	PPB	12/29/2023 18:32	SMPL
O5815-17	E0ZF7	3304	0.5504 -	SFAM01.1	PPB	12/29/2023 18:35	SMPL
O5815-18	E0ZF8	11550	2.1605 -	SFAM01.1	PPB	12/29/2023 18:42	SMPL
O5815-20	E0ZF8D	12242	2.2957 -	SFAM01.1	PPB	12/29/2023 18:44	SMPL
O5815-21	E0ZF8S	22327	4.2649 -	SFAM01.1	PPB	12/29/2023 18:46	SMPL
PB158172BL	PBS172	396	-0.0175 -	SFAM01.1	PPB	12/29/2023 18:49	SMPL
O5816-01	E0Z26	1385	0.1757 -	SFAM01.1	PPB	12/29/2023 18:51	SMPL
O5816-02	E0Z27	7495	1.3687 -	SFAM01.1	PPB	12/29/2023 18:53	SMPL
O5816-03	E0Z28	4718	0.8265 -	SFAM01.1	PPB	12/29/2023 18:56	SMPL
O5816-04	E0Z29	3838	0.6546 -	SFAM01.1	PPB	12/29/2023 18:58	SMPL
O5816-05	E0ZG6	1917	0.2795 -	SFAM01.1	PPB	12/29/2023 19:00	SMPL
O5816-06	E0ZG8	975	0.0956 -	SFAM01.1	PPB	12/29/2023 19:02	SMPL
O5816-07	E0ZG9	732	0.0481 -	SFAM01.1	PPB	12/29/2023 19:05	SMPL
O5816-08	E0ZH0	664	0.0349 -	SFAM01.1	PPB	12/29/2023 19:07	SMPL
O5816-09	E0Z30	9735	1.8061 -	SFAM01.1	PPB	12/29/2023 19:09	SMPL
O5816-10	E0Z46	6950	1.2623 -	SFAM01.1	PPB	12/29/2023 19:12	SMPL
O5816-11	E0Z47	10485	1.9526 -	SFAM01.1	PPB	12/29/2023 19:14	SMPL
O5816-12	E0Z48	14895	2.8137 -	SFAM01.1	PPB	12/29/2023 19:16	SMPL
O5816-13	E0Z49	22527	4.304 -	SFAM01.1	PPB	12/29/2023 19:18	SMPL
O5816-14	E0Z50	16751	3.1761 -	SFAM01.1	PPB	12/29/2023 19:21	SMPL
O5816-15	E0ZG7	2308	0.3559 -	SFAM01.1	PPB	12/29/2023 19:23	SMPL
CCV094	CCV094	26428	5.0657 -	SFAM01.1	PPB	12/29/2023 19:25	SMPL
CCB094	CCB094	441	-0.0087 -	SFAM01.1	PPB	12/29/2023 19:28	SMPL
O5816-16	E0ZJ1	1734	0.2438 -	SFAM01.1	PPB	12/29/2023 19:30	SMPL
O5816-17	E0ZJ2	1169	0.1335 -	SFAM01.1	PPB	12/29/2023 19:32	SMPL
O5816-18	E0ZJ3	1749	0.2467 -	SFAM01.1	PPB	12/29/2023 19:34	SMPL
O5816-19	E0ZJ4	278	-0.0405 -	SFAM01.1	PPB	12/29/2023 19:37	SMPL
O5816-20	E0ZJ5	331	-0.0302 -	SFAM01.1	PPB	12/29/2023 19:39	SMPL
O5816-21	E0ZJ5D	361	-0.0243 -	SFAM01.1	PPB	12/29/2023 19:41	SMPL
O5816-22	E0ZJ5S	14838	2.8026 -	SFAM01.1	PPB	12/29/2023 19:44	SMPL
CCV095	CCV095	26429	5.0659 -	SFAM01.1	PPB	12/29/2023 19:46	SMPL
CCB095	CCB095	432	-0.0104 -	SFAM01.1	PPB	12/29/2023 19:48	SMPL