

LB128996


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58570



A= 0.0000e+000

B= 1.6749e-004

C= -9.7682e-002

Rho= 0.9988178

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o D
0.0	0.000	-0.086	-0.086	69	0.000	69					
0.2	0.200	0.187	-0.013	1699	0.0 %	1699					-7
2.5	2.500	2.444	-0.056	15176	0.0 %	15176					-2
5.0	5.000	5.198	0.198	31617	0.0 %	31617					4
7.5	7.500	7.746	0.246	46829	0.0 %	46829					3
10.0	10.000	9.712	-0.288	58570	0.0 %	58570					-3

LB128996 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Jan 2024 08:42:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	69	-	0.00007471B	PPB		02 Jan 2024 08:44:51	S	Std
0.2 - 1	50.2	1699	-	0.20007471B	PPB		02 Jan 2024 08:47:08	S	Std
2.5 - 1	52.5	15176	-	2.50007471B	PPB		02 Jan 2024 08:49:26	S	Std
5.0 - 1	55	31617	-	5.00007471B	PPB		02 Jan 2024 08:51:43	S	Std
7.5 - 1	57.5	46829	-	7.50007471B	PPB		02 Jan 2024 08:54:02	S	Std
10.0 - 1	59.5	58570	-	10.00007471B	PPB		02 Jan 2024 09:10:24	S	Std
ICV47 - 1	ICV47	24781	4.0528	-7471B	PPB		02 Jan 2024 09:13:34	U	SMPL
ICB47 - 1	ICB47	214	-0.0616	-7471B	PPB		02 Jan 2024 09:18:37	U	SMPL
CCV01 - 1	CCV01	32226	5.2997	-7471B	PPB		02 Jan 2024 09:20:52	U	SMPL
CCB01 - 1	CCB01	192	-0.0655	-7471B	PPB		02 Jan 2024 09:25:51	U	SMPL
CRA - 1	CRA	1633	0.1758	-7471B	PPB		02 Jan 2024 09:28:06	U	SMPL
HighStd - 1	HighStd	60014	9.9538	-7471B	PPB		02 Jan 2024 09:30:21	U	SMPL
ChkStd - 1	ChkStd	44497	7.3549	-7471B	PPB		02 Jan 2024 09:32:36	U	SMPL
PB158175BL - 1	PB158175BL	283	-0.0503	-7471B	PPB		02 Jan 2024 09:37:52	U	SMPL
PB158175BS - 1	PB158175BS	24752	4.0479	-7471B	PPB		02 Jan 2024 09:40:07	U	SMPL
O5903-01 - 1	GHL-SD-42	4687	0.6873	-7471B	PPB		02 Jan 2024 09:42:24	U	SMPL
O5903-02 - 1	GHL-SS-18-0-2	176642	29.4873	-7471B	PPB		02 Jan 2024 09:44:46	U	SMPL
O5903-03 - 1	GHL-SS-18-2-4	306441	51.2268	-7471B	PPB		02 Jan 2024 09:47:05	U	SMPL
O5903-04 - 1	GHL-SS-19-0-2	24718	4.0422	-7471B	PPB		02 Jan 2024 09:49:51	U	SMPL
O5903-05 - 1	GHL-SS-19-2-4	116208	19.3655	-7471B	PPB		02 Jan 2024 09:52:35	U	SMPL
CCV02 - 1	CCV02	29560	4.8532	-7471B	PPB		02 Jan 2024 09:57:58	U	SMPL
CCB02 - 1	CCB02	357	-0.0379	-7471B	PPB		02 Jan 2024 10:02:55	U	SMPL
O5903-06 - 1	GHL-SS-20	7061	1.0849	-7471B	PPB		02 Jan 2024 10:05:10	U	SMPL
O5903-07 - 1	GHL-SS-21	14652	2.3563	-7471B	PPB		02 Jan 2024 10:07:26	U	SMPL
O5903-08 - 1	GHL-SS-22	48211	7.9770	-7471B	PPB		02 Jan 2024 10:09:44	U	SMPL
O5903-09 - 1	GHL-SS-23	8481	1.3228	-7471B	PPB		02 Jan 2024 10:12:05	U	SMPL
O5903-10 - 1	GHL-SS-24	6931	1.0632	-7471B	PPB		02 Jan 2024 10:14:33	U	SMPL
O5903-11 - 1	GHL-SS-26	6605	1.0086	-7471B	PPB		02 Jan 2024 10:16:52	U	SMPL
O5903-12 - 1	GHL-SS-27	17538	2.8397	-7471B	PPB		02 Jan 2024 10:19:11	U	SMPL
O5903-13 - 1	GHL-SS-28	3840	0.5455	-7471B	PPB		02 Jan 2024 10:21:31	U	SMPL
O5903-14 - 1	GHL-SS-29	5254	0.7823	-7471B	PPB		02 Jan 2024 10:23:51	U	SMPL
O5903-15 - 1	GHL-SS-30	7289	1.1231	-7471B	PPB		02 Jan 2024 10:26:08	U	SMPL
CCV03 - 1	CCV03	31015	5.0969	-7471B	PPB		02 Jan 2024 10:28:26	U	SMPL
CCB03 - 1	CCB03	263	-0.0536	-7471B	PPB		02 Jan 2024 10:33:58	U	SMPL
O5903-16 - 1	GHL-SS-31	7531	1.1637	-7471B	PPB		02 Jan 2024 10:36:14	U	SMPL
O5903-17 - 1	GHL-SS-32	9836	1.5500	-7471B	PPB		02 Jan 2024 10:38:29	U	SMPL
O5903-18 - 1	GHL-SS-33	11898	1.8951	-7471B	PPB		02 Jan 2024 10:40:47	U	SMPL
O5903-19 - 1	GHL-SS-34	16896	2.7322	-7471B	PPB		02 Jan 2024 10:43:05	U	SMPL
O5903-20 - 1	GHL-SS-39	2158	0.2638	-7471B	PPB		02 Jan 2024 10:45:25	U	SMPL
O5903-21 - 1	GHL-SS-40	2585	0.3353	-7471B	PPB		02 Jan 2024 10:47:44	U	SMPL
O5903-22 - 1	GHL-SS-41	2843	0.3785	-7471B	PPB		02 Jan 2024 10:50:01	U	SMPL
O5903-23 - 1	GHL-SS-43	2663	0.3483	-7471B	PPB		02 Jan 2024 10:52:19	U	SMPL
O5903-23DUP - 1	GHL-SS-43DUP	2918	0.3910	-7471B	PPB		02 Jan 2024 10:58:05	U	SMPL
O5903-23MS - 1	GHL-SS-43MS	31130	5.1161	-7471B	PPB		02 Jan 2024 11:00:21	U	SMPL
CCV04 - 1	CCV04	28389	4.6571	-7471B	PPB		02 Jan 2024 11:05:23	U	SMPL
CCB04 - 1	CCB04	170	-0.0692	-7471B	PPB		02 Jan 2024 11:07:40	U	SMPL
O5903-23MSD - 1	GHL-SS-43MSD	32527	5.3501	-7471B	PPB		02 Jan 2024 11:10:02	U	SMPL
PB158176BL - 1	PB158176BL	198	-0.0645	-7471B	PPB		02 Jan 2024 11:12:18	U	SMPL
PB158176BS - 1	PB158176BS	23640	3.8617	-7471B	PPB		02 Jan 2024 11:14:43	U	SMPL
O5904-01 - 1	GHL-SS-01	30107	4.9448	-7471B	PPB		02 Jan 2024 11:16:59	U	SMPL
O5904-02 - 1	GHL-SS-02	12901	2.0631	-7471B	PPB		02 Jan 2024 11:19:20	U	SMPL
O5904-03 - 1	GHL-SS-13	50808	8.4119	-7471B	PPB		02 Jan 2024 11:21:47	U	SMPL
O5904-04 - 1	GHL-SS-03	12469	1.9907	-7471B	PPB		02 Jan 2024 11:24:07	U	SMPL
O5904-05 - 1	GHL-SS-04	13464	2.1573	-7471B	PPB		02 Jan 2024 11:26:37	U	SMPL
O5904-06 - 1	GHL-SS-05	16317	2.6352	-7471B	PPB		02 Jan 2024 11:28:57	U	SMPL
O5904-07 - 1	GHL-SS-06-0-2	5724	0.8610	-7471B	PPB		02 Jan 2024 11:31:16	U	SMPL
CCV05 - 1	CCV05	28699	4.7090	-7471B	PPB		02 Jan 2024 11:36:28	U	SMPL
CCB05 - 1	CCB05	281	-0.0506	-7471B	PPB		02 Jan 2024 11:41:25	U	SMPL
O5904-08 - 1	GHL-SS-06-2-4	15521	2.5019	-7471B	PPB		02 Jan 2024 11:43:40	U	SMPL
O5904-09 - 1	GHL-SS-08	15050	2.4230	-7471B	PPB		02 Jan 2024 11:45:55	U	SMPL
O5904-10 - 1	GHL-SS-09	7303	1.1255	-7471B	PPB		02 Jan 2024 11:48:14	U	SMPL
O5904-11 - 1	GHL-SS-10-0-2	107842	17.9643	-7471B	PPB		02 Jan 2024 11:50:35	U	SMPL
O5904-12 - 1	GHL-SS-10-2-4	6580	1.0044	-7471B	PPB		02 Jan 2024 11:52:55	U	SMPL
O5904-13 - 1	GHL-SS-11	6759	1.0344	-7471B	PPB		02 Jan 2024 11:55:31	U	SMPL
O5904-14 - 1	GHL-SS-12	52294	8.6608	-7471B	PPB		02 Jan 2024 11:57:50	U	SMPL
O5904-15 - 1	GHL-SS-14-0-2	9932	1.5658	-7471B	PPB		02 Jan 2024 12:00:09	U	SMPL
O5904-16 - 1	GHL-SS-15-0-2	11910	1.8971	-7471B	PPB		02 Jan 2024 12:02:38	U	SMPL
O5904-17 - 1	GHL-SS-15-2-4	5706	0.8580	-7471B	PPB		02 Jan 2024 12:04:58	U	SMPL
CCV06 - 1	CCV06	30829	5.0657	-7471B	PPB		02 Jan 2024 12:07:17	U	SMPL
CCB06 - 1	CCB06	471	-0.0188	-7471B	PPB		02 Jan 2024 12:09:36	U	SMPL
O5904-18 - 1	GHL-SS-16-0-2	8911	1.3948	-7471B	PPB		02 Jan 2024 12:12:00	U	SMPL
O5904-19 - 1	GHL-SS-16-2-4	7094	1.0905	-7471B	PPB		02 Jan 2024 12:14:16	U	SMPL
O5904-20 - 1	GHL-SS-17	79654	13.2432	-7471B	PPB		02 Jan 2024 12:16:50	U	SMPL
O5904-20DUP - 1	GHL-SS-17DUP	83650	13.9125	-7471B	PPB		02 Jan 2024 12:24:09	U	SMPL
O5904-20MS - 1	GHL-SS-17MS	119424	19.9041	-7471B	PPB		02 Jan 2024 12:26:41	U	SMPL
O5904-20MSD - 1	GHL-SS-17MSD	118706	19.7839	-7471B	PPB		02 Jan 2024 12:32:56	U	SMPL
PB158177BL - 1	PB158177BL	105	-0.0801	-7471B	PPB		02 Jan 2024 12:38:16	U	SMPL
PB158177BS - 1	PB158177BS	25349	4.1479	-7471B	PPB		02 Jan 2024 12:40:32	U	SMPL
O5902-21 - 1	GHL-SD-52-D	6920	1.0613	-7471B	PPB		02 Jan 2024 12:42:49	U	SMPL
O5902-21DUP - 1	GHL-SD-52-DDUP	6910	1.0596	-7471B	PPB		02 Jan 2024 12:45:10	U	SMPL
CCV07 - 1	CCV07	29968	4.9215	-7471B	PPB		02 Jan 2024 12:49:50	U	SMPL

LB128996 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Jan 2024 08:42:39

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB07 - 1	CCB07	30	-0.0927		-7471B	PPB	02 Jan 2024 12:54:57	U	SMPL
O5902-21MS - 1	GHL-SD-52-DMS	31655	5.2041		-7471B	PPB	02 Jan 2024 12:57:13	U	SMPL
O5902-21MSD - 1	GHL-SD-52-DMSD	29677	4.8728		-7471B	PPB	02 Jan 2024 12:59:29	U	SMPL
O5902-22 - 1	GHL-SD-35a	3242	0.4453		-7471B	PPB	02 Jan 2024 13:01:52	U	SMPL
O5902-23 - 1	GHL-SD-36a	8142	1.2660		-7471B	PPB	02 Jan 2024 13:04:16	U	SMPL
O5902-24 - 1	GHL-SD-37a	4504	0.6567		-7471B	PPB	02 Jan 2024 13:06:35	U	SMPL
O6015-01 - 1	MR-317-0-2	24735	4.0451		-7471B	PPB	02 Jan 2024 13:08:54	U	SMPL
O6015-03 - 1	MR-317-0-2D	21924	3.5743		-7471B	PPB	02 Jan 2024 13:11:12	U	SMPL
O6015-05 - 1	MR-323-0-2	7983	1.2394		-7471B	PPB	02 Jan 2024 13:13:35	U	SMPL
O6015-07 - 1	MR-323-6-8	7799	1.2085		-7471B	PPB	02 Jan 2024 13:15:58	U	SMPL
CCV08 - 1	CCV08	31568	5.1895		-7471B	PPB	02 Jan 2024 13:18:17	U	SMPL
CCB08 - 1	CCB08	380	-0.0340		-7471B	PPB	02 Jan 2024 13:20:36	U	SMPL
O5903-23LX5 - 1		624	0.0068		-7471B	PPB	02 Jan 2024 13:22:59	U	SMPL
O5903-23A - 1		32111	5.2805		-7471B	PPB	02 Jan 2024 13:34:33	U	SMPL
O5904-20LX5 - 1		12504	1.9966		-7471B	PPB	02 Jan 2024 13:36:49	U	SMPL
O5904-20A - 1		119032	19.8385		-7471B	PPB	02 Jan 2024 13:39:14	U	SMPL
O5902-21LX5 - 1		585	0.0003		-7471B	PPB	02 Jan 2024 13:41:36	U	SMPL
O5902-21A - 1		31018	5.0974		-7471B	PPB	02 Jan 2024 13:44:14	U	SMPL
O5903-02DLX5 - 1		33466	5.5074		-7471B	PPB	02 Jan 2024 13:55:27	U	SMPL
O5903-03DLX10 - 1		29507	4.8443		-7471B	PPB	02 Jan 2024 14:00:08	U	SMPL
O5903-05DLX5 - 1		21472	3.4986		-7471B	PPB	02 Jan 2024 14:02:30	U	SMPL
O5904-11DLX5 - 1		23684	3.8690		-7471B	PPB	02 Jan 2024 14:07:13	U	SMPL
CCV09 - 1	CCV09	31366	5.1557		-7471B	PPB	02 Jan 2024 14:09:37	U	SMPL
CCB09 - 1	CCB09	283	-0.0503		-7471B	PPB	02 Jan 2024 14:11:54	U	SMPL
O5904-20DLX5 - 1		16284	2.6297		-7471B	PPB	02 Jan 2024 14:14:14	U	SMPL
O5904-20DUPDLX5 - 1		17844	2.8909		-7471B	PPB	02 Jan 2024 14:16:35	U	SMPL
O5904-20MSDLX5 - 1		24218	3.9585		-7471B	PPB	02 Jan 2024 14:21:16	U	SMPL
O5904-20MSDDLX5 - 1		25905	4.2410		-7471B	PPB	02 Jan 2024 14:23:33	U	SMPL
O5904-20LDLX25 - 1		1819	0.2070		-7471B	PPB	02 Jan 2024 14:29:28	U	SMPL
O5904-20ADLX5 - 1		18311	2.9691		-7471B	PPB	02 Jan 2024 14:34:25	U	SMPL
CCV10 - 1	CCV10	31211	5.1297		-7471B	PPB			
CCB10 - 1	CCB10	28	-0.0930		-7471B	PPB			