



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129000

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	01/02/2024	13:15
2	O6038-07		1	12	100.00	>212.0	>212.0	01/02/2024	13:45
3	O6038-07DUP		1	12	100.00	>212.0	>212.0	01/02/2024	14:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 129000

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint 01/02/24

WorkList ID : 176884

Department : Wet-Chemistry

Date : 01-02-2024 11:35:46

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
O6038-07	EQ-TANK	Water	Flash Point	Cool 4 deg C	PSEG03	O31	12/28/2023	1010B

Date/Time 01/02/2024 13:05
Raw Sample Received by: N.P (W.L)
Raw Sample Relinquished by: N.P (W.C)

Date/Time 01/02/2024 (15:00)
Raw Sample Received by: N.P (W.C)
Raw Sample Relinquished by: N.P (W.C)