



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129002

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/02/2024	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	01/02/2024	16:02
3	CAL3	1	Water	NA	NA	20.3	10.02	01/02/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.02	01/02/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	01/02/2024	16:11
6	P1003-02	1	Solid	20.02	20	22.9	8.30	01/02/2024	16:20
7	P1003-02DUP	1	Solid	20.02	20	22.8	8.31	01/02/2024	16:21
8	CCV2	1	Water	NA	NA	20.3	12.02	01/02/2024	16:30

WORKLIST(Hardcopy Internal Chain)

VB 29002

WorkList Name : CORROSIONITY P1003

WorkList ID : 176904

Department : Wet-Chemistry

Date : 01-02-2024 15:41:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1003-02	72-11978	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	01/02/2024	9045D

Date/Time 01/02/24 15:50

Raw Sample Received by: VB (CPC)

Raw Sample Relinquished by: JTC (SM)

Date/Time 01/02/24 17:30

Raw Sample Received by: JTC (SM)

Raw Sample Relinquished by: JB (CPC)