



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129008

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	01/03/2024	08:45
2	P1001-01		1	12	100.00	>212.0	>212.0	01/03/2024	09:15
3	P1001-01DUP		1	12	100.00	>212.0	>212.0	01/03/2024	09:45
4	P1001-02		1	13	100.00	>212.0	>212.0	01/03/2024	10:15
5	P1001-03		1	14	100.00	>212.0	>212.0	01/03/2024	10:45
6	P1001-04		1	14	100.00	>212.0	>212.0	01/03/2024	11:15
7	P1002-01		1	14	100.00	>212.0	>212.0	01/03/2024	11:45
8	P1002-02		1	12	100.00	>212.0	>212.0	01/03/2024	12:15
9	P1002-03		1	14	100.00	>212.0	>212.0	01/03/2024	12:45
10	P1002-04		1	13	100.00	>212.0	>212.0	01/03/2024	13:15
11	P1002-05		1	12	100.00	>212.0	>212.0	01/03/2024	13:45
12	P1002-06		1	13	100.00	>212.0	>212.0	01/03/2024	14:15
13	P1002-07		1	14	100.00	>212.0	>212.0	01/03/2024	14:45
14	P1002-08		1	13	100.00	>212.0	>212.0	01/03/2024	15:15
15	P1002-09		1	13	100.00	>212.0	>212.0	01/03/2024	15:45
16	P1002-10		1	13	100.00	>212.0	>212.0	01/03/2024	16:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB.129008

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint01/03/2024

WorkList ID : 176913

Department : Wet-Chemistry

Date : 01-03-2024 08:24:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1001-01	ORA-1205	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1001-02	ORA-1206	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1001-03	ORA-1207	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1001-04	ORA-1208	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-01	1209	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-02	1210	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-03	1211	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-04	1212	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-05	1213	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-06	1214	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-07	1215	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-08	1216	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-09	1217	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B
P1002-10	1218	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B

Date/Time 01/03/2024 8:30

Raw Sample Received by: N.P. W.C.

Raw Sample Relinquished by: M. G. C.

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: