

LB129009


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

42625



A= 0.0000e+000

B= 2.3342e-004

C= -4.3858e-002

Rho= 0.9996293

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.049	0.049	398	0.000	398	0				~
0.2	0.200	0.222	0.022	1139	0.0 %	1139					11
2.5	2.500	2.374	-0.126	10357	0.0 %	10357					-5
5.0	5.000	4.971	-0.029	21485	0.0 %	21485					1
7.5	7.500	7.679	0.179	33085	0.0 %	33085					2
10.0	10.000	9.906	-0.094	42625	0.0 %	42625					~1

LB129009 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jan 2024 09:48:02

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	398	-	0.00007470A	PPB		03 Jan 2024 10:04:35	S	Std
0.2 - 1	50.2	1139	-	0.20007470A	PPB		03 Jan 2024 10:06:52	S	Std
2.5 - 1	52.7	10357	-	2.50007470A	PPB		03 Jan 2024 10:09:08	S	Std
5.0 - 1	55	21485	-	5.00007470A	PPB		03 Jan 2024 10:11:25	S	Std
7.5 - 1	57.5	33085	-	7.50007470A	PPB		03 Jan 2024 10:16:30	S	Std
10.0 - 1	59	42625	-	10.00007470A	PPB		03 Jan 2024 10:21:37	S	Std
ICV48 - 1	ICV48	17688	4.0848	-7470A	PPB		03 Jan 2024 10:26:57	U	SMPL
ICB48 - 1	ICB48	81	-0.0250	-7470A	PPB		03 Jan 2024 10:29:14	U	SMPL
CCV11 - 1	CCV11	21486	4.9713	-7470A	PPB		03 Jan 2024 10:31:32	U	SMPL
CCB11 - 1	CCB11	138	-0.0116	-7470A	PPB		03 Jan 2024 10:33:47	U	SMPL
CRA - 1	CRA	1030	0.1966	-7470A	PPB		03 Jan 2024 10:36:05	U	SMPL
HighStd - 1	HighStd	42866	9.9625	-7470A	PPB		03 Jan 2024 10:38:20	U	SMPL
ChkStd - 1	ChkStd	31781	7.3743	-7470A	PPB		03 Jan 2024 10:40:35	U	SMPL
PB158165BL - 1	PBW	253	0.0152	-7470A	PPB		03 Jan 2024 10:45:47	U	SMPL
PB158165BS - 1	LCSW	19466	4.4980	-7470A	PPB		03 Jan 2024 10:48:03	U	SMPL
O6031-02 - 1	MR-321-0-2	-122	-0.0723	-7470A	PPB		03 Jan 2024 10:50:20	U	SMPL
O6031-04 - 1	MR-321-10-12	166	-0.0051	-7470A	PPB		03 Jan 2024 10:52:38	U	SMPL
O6037-01 - 1	BLDG-39-STRUCTURAL-STEEL-PAINT-REMOVAL	24	-0.0383	-7470A	PPB		03 Jan 2024 10:54:56	U	SMPL
O6038-02 - 1	T-402	293	0.0245	-7470A	PPB		03 Jan 2024 10:57:15	U	SMPL
O6038-04 - 1	T-401	18	-0.0397	-7470A	PPB		03 Jan 2024 10:59:30	U	SMPL
CCV12 - 1	CCV12	21515	4.9781	-7470A	PPB		03 Jan 2024 11:01:47	U	SMPL
CCB12 - 1	CCB12	131	-0.0133	-7470A	PPB		03 Jan 2024 11:04:01	U	SMPL
O6038-06 - 1	T-803	114	-0.0172	-7470A	PPB		03 Jan 2024 11:06:19	U	SMPL
O6038-10 - 1	T-703	204	0.0038	-7470A	PPB		03 Jan 2024 11:08:36	U	SMPL
O6040-02 - 1	EMS-38-WC-1	175	-0.0030	-7470A	PPB		03 Jan 2024 11:10:51	U	SMPL
O6040-02DUP - 1	EMS-38-WC-1DUP	470	0.0658	-7470A	PPB		03 Jan 2024 11:15:45	U	SMPL
O6040-02MS - 1	EMS-38-WC-1MS	20486	4.7379	-7470A	PPB		03 Jan 2024 11:18:01	U	SMPL
O6040-02MSD - 1	EMS-38-WC-1MSD	19502	4.5082	-7470A	PPB		03 Jan 2024 11:20:18	U	SMPL
O6045-02 - 1	MR-300-0-2	-59	-0.0576	-7470A	PPB		03 Jan 2024 11:22:37	U	SMPL
O6045-04 - 1	MR-300-2.0-2.5	125	-0.0147	-7470A	PPB		03 Jan 2024 11:24:56	U	SMPL
CCV13 - 1	CCV13	20953	4.8469	-7470A	PPB		03 Jan 2024 11:29:48	U	SMPL
CCB13 - 1	CCB13	129	-0.0137	-7470A	PPB		03 Jan 2024 11:32:05	U	SMPL
PB158107TB - 1	PB158107TB	76	-0.0257	-7470A	PPB		03 Jan 2024 11:34:23	U	SMPL
PB158150TB - 1	PB158150TB	147	-0.0095	-7470A	PPB		03 Jan 2024 11:36:39	U	SMPL
O6040-02LX5 - 1		30	-0.0369	-7470A	PPB		03 Jan 2024 11:38:55	U	SMPL
O6040-02A - 1		18872	4.3612	-7470A	PPB		03 Jan 2024 11:41:10	U	SMPL
O5831-06 - 1	CHRT-20017	100	-0.0205	-7470A	PPB		03 Jan 2024 11:49:14	U	SMPL
CCV14 - 1	CCV14	22822	5.2832	-7470A	PPB		03 Jan 2024 11:51:29	U	SMPL
CCB14 - 1	CCB14	118	-0.0163	-7470A	PPB		03 Jan 2024 11:53:45	U	SMPL