


LB129010

7471B

INSTRUMENT ID: CVL

μ Abs.:

7030E



Linear

A= 0.0000e+000

B= 1.4168e-004

C= 3.3926e-002

Rho= 0.9983636

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.040	0.040	44	0.000	44					
0.2	0.200	0.186	-0.014	1076	0.0 %	1076					
2.5	2.500	2.304	-0.196	16023	0.0 %	16023					
5.0	5.000	5.413	0.413	37966	0.0 %	37966					
7.5	7.500	7.262	-0.238	51020	0.0 %	51020					
10.0	10.000	9.995	-0.005	70306	0.0 %	70306					

57
18
28
22
0

LB129010 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jan 2024 13:52:51

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	44	-	0.0000	7471B	PPB	03 Jan 2024 13:56:41	S	Std
0.2 - 1	50.2	1076	-	0.2000	7471B	PPB	03 Jan 2024 13:58:58	S	Std
2.5 - 1	52.5	16023	-	2.5000	7471B	PPB	03 Jan 2024 14:01:15	S	Std
5.0 - 1	55	37966	-	5.0000	7471B	PPB	03 Jan 2024 14:03:32	S	Std
7.5 - 1	57.5	51020	-	7.5000	7471B	PPB	03 Jan 2024 14:05:55	S	Std
10.0 - 1	5/11	70306	-	10.0000	7471B	PPB	03 Jan 2024 14:13:58	S	Std
ICV49 - 1	ICV49	26103	3.7321	-	7471B	PPB	03 Jan 2024 14:22:06	U	SMPL
ICB49 - 1	ICB49	-348	-0.0154	-	7471B	PPB	03 Jan 2024 14:24:23	U	SMPL
CCV15 - 1	CCV15	38011	5.4191	-	7471B	PPB	03 Jan 2024 14:26:49	U	SMPL
CCB15 - 1	CCB15	-414	-0.0247	-	7471B	PPB	03 Jan 2024 14:29:04	U	SMPL
CRA - 1	CRA	994	0.1748	-	7471B	PPB	03 Jan 2024 14:31:30	U	SMPL
HighStd - 1	HighStd	70686	10.0484	-	7471B	PPB	03 Jan 2024 14:33:46	U	SMPL
ChkStd - 1	ChkStd	48533	6.9098	-	7471B	PPB	03 Jan 2024 14:36:01	U	SMPL
PB158220BL - 1	PBS	-220	0.0028	-	7471B	PPB	03 Jan 2024 14:38:33	U	SMPL
PB158220BS - 1	LCSS	23468	3.3588	-	7471B	PPB	03 Jan 2024 14:43:50	U	SMPL
O6038-01 - 1	T-402	13427	1.9362	-	7471B	PPB	03 Jan 2024 14:46:06	U	SMPL
O6038-03 - 1	T-401	5987	0.8621	-	7471B	PPB	03 Jan 2024 14:48:32	U	SMPL
O6038-05 - 1	T-803	2455	0.3817	-	7471B	PPB	03 Jan 2024 14:50:52	U	SMPL
O6038-09 - 1	T-703	2143	0.3375	-	7471B	PPB	03 Jan 2024 14:53:12	U	SMPL
O6040-01 - 1	EMS-38-WC-1	46389	6.6061	-	7471B	PPB	03 Jan 2024 14:55:29	U	SMPL
CCV16 - 1	CCV16	36808	5.2487	-	7471B	PPB	03 Jan 2024 14:57:45	U	SMPL
CCB16 - 1	CCB16	-279	-0.0056	-	7471B	PPB	03 Jan 2024 15:00:14	U	SMPL
O6041-01 - 1	BU-02-12282023	13207	1.9050	-	7471B	PPB	03 Jan 2024 15:02:42	U	SMPL
O6045-01 - 1	MR-300-0-2	24057	3.4422	-	7471B	PPB	03 Jan 2024 15:04:57	U	SMPL
O6045-03 - 1	MR-300-2.0-2.5	19863	2.8480	-	7471B	PPB	03 Jan 2024 15:11:52	U	SMPL
O6045-03DUP - 1	MR-300-2.0-2.5DUP	19392	2.7813	-	7471B	PPB	03 Jan 2024 15:14:12	U	SMPL
O6045-03MS - 1	MR-300-2.0-2.5MS	41581	5.9249	-	7471B	PPB	03 Jan 2024 15:16:33	U	SMPL
O6045-03MSD - 1	MR-300-2.0-2.5MSD	42763	6.0924	-	7471B	PPB	03 Jan 2024 15:18:52	U	SMPL
O6045-03LX5 - 1		4213	0.6308	-	7471B	PPB	03 Jan 2024 15:21:18	U	SMPL
O6045-03A - 1		46158	6.5734	-	7471B	PPB	03 Jan 2024 15:23:46	U	SMPL
CCV17 - 1	CCV17	37597	5.3605	-	7471B	PPB	03 Jan 2024 15:31:24	U	SMPL
CCB17 - 1	CCB17	-281	-0.0059	-	7471B	PPB	03 Jan 2024 15:33:41	U	SMPL