

LB129019


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

59318



A= 0.0000e+000

B= 1.6883e-004 slope

C= -1.3704e-002 y-intercept

Rho= 0.9997660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.00	0.000	0.002	0.002	92	0.000	92					40
0.05	0.050					442	0				=
0.20	0.200	0.175	-0.025	1119	0.0 %	1119					-13
2.50	2.500	2.615	0.115	15570	0.0 %	15570					5
5.00	5.000	4.853	-0.147	28827	0.0 %	28827					-3
7.50	7.500	7.553	0.053	44820	0.0 %	44820					1
10.0	10.000	10.001	0.001	59318	0.0 %	59318					0

LB129019
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	92	-		0 SFAM01.1	PPB	1/4/2024 13:53	Std
	0.2 S01	1119	-		0.2 SFAM01.1	PPB	1/4/2024 13:57	Std
	2.5 S02	15570	-		2.5 SFAM01.1	PPB	1/4/2024 14:00	Std
	5 S03	28827	-		5 SFAM01.1	PPB	1/4/2024 14:05	Std
	7.5 S04	44820	-		7.5 SFAM01.1	PPB	1/4/2024 14:10	Std
	10 S05	59318	-		10 SFAM01.1	PPB	1/4/2024 14:14	Std
ICV051	ICV051	23032	3.8749 -		SFAM01.1	PPB	1/4/2024 14:18	SMPL
ICB051	ICB051	-111	-0.0324 -		SFAM01.1	PPB	1/4/2024 14:21	SMPL
CCV023	CCV023	28733	4.8374 -		SFAM01.1	PPB	1/4/2024 14:26	SMPL
CCB023	CCB023	-170	-0.0424 -		SFAM01.1	PPB	1/4/2024 14:28	SMPL
PB158245BL	PBW245	108	0.0045 -		SFAM01.1	PPB	1/4/2024 14:30	SMPL
O5867-01	GCET6	240	0.0268 -		SFAM01.1	PPB	1/4/2024 14:32	SMPL
O5867-02	GCET6D	236	0.0261 -		SFAM01.1	PPB	1/4/2024 14:35	SMPL
O5867-03	GCET6S	5989	0.9974 -		SFAM01.1	PPB	1/4/2024 14:42	SMPL
O5867-04	GCET7	36	-0.0076 -		SFAM01.1	PPB	1/4/2024 14:44	SMPL
O5867-05	GCET8	116	0.0059 -		SFAM01.1	PPB	1/4/2024 14:46	SMPL
O5867-06	GCET9	130	0.0082 -		SFAM01.1	PPB	1/4/2024 14:49	SMPL
O5867-07	GCEW0	143	0.0104 -		SFAM01.1	PPB	1/4/2024 14:51	SMPL
O5867-08	GCEW1	152	0.012 -		SFAM01.1	PPB	1/4/2024 14:53	SMPL
O5867-09	GCEW2	107	0.0044 -		SFAM01.1	PPB	1/4/2024 14:56	SMPL
O5867-10	GCEW3	310	0.0386 -		SFAM01.1	PPB	1/4/2024 14:58	SMPL
O5867-11	GCF03	368	0.0484 -		SFAM01.1	PPB	1/4/2024 15:00	SMPL
O5867-12	GCF04	606	0.0886 -		SFAM01.1	PPB	1/4/2024 15:02	SMPL
O5867-13	GCF05	1064	0.1659 -		SFAM01.1	PPB	1/4/2024 15:05	SMPL
CCV024	CCV024	28068	4.7251 -		SFAM01.1	PPB	1/4/2024 15:07	SMPL
CCB024	CCB024	-161	-0.0409 -		SFAM01.1	PPB	1/4/2024 15:09	SMPL