



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB129022
Analysis Date: 01/04/2024
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 01/04/2024
Extraction IN Time: 12:35
Extraction OUT Time: 13:30
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB129022BL	LB129022BL	WATER	1.3	1000	100	3.1502	3.1502	3.02	3.1503	3.1503	0.0001	0.1
2	LB129022BS	LB129022BS	WATER	1.3	1000	100	3.1988	3.1988	3.01	3.2157	3.2157	0.0169	16.9
3	LB129022BSD	LB129022BSD	WATER	1.3	1000	100	3.0225	3.0225	3.03	3.0395	3.0395	0.0170	17
4	O6038-07	EQ-TANK	WATER	1.6	1000	100	3.0486	3.0486	3.04	3.0490	3.0490	0.0004	0.4
5	P1008-01	402	WATER	1.3	1000	100	3.0167	3.0167	3.03	3.0168	3.0168	0.0001	0.1
6	P1014-01	GRAB	WATER	1.6	1000	100	3.0081	3.0081	3.02	3.0687	3.0687	0.0606	60.6



Chemical Name	Chemical Lot #
HEXANE	W3042
pH Paper 0-14	M4909
Sodium Sulfate	EP2429
1:1 HCL	WP105958
Silica Gel	W2839
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP105959
LCSWD	5.00 ML	WP105960
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0018</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1 :	<u>15:11</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>14:20</u>		
Bal Check Time:	<u>12:55</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1:	<u>15:45</u>
			Out Time1:	<u>15:10</u>		

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2 :	<u>17:00</u>
1.0000 gram Balance:	<u>1.0003</u>	(0.9950-1.0050)	In Time2:	<u>16:10</u>		
Bal Check Time:	<u>17:33</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2:	<u>17:30</u>
			Out Time2:	<u>16:35</u>		