



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129024

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	01/04/2024	09:00
2	P1002-12		1	6	28.00	82.4	81.4	01/04/2024	09:30
3	P1002-12DUP		1	6	27.00	80.6	79.6	01/04/2024	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 129007
LB 129024

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint s01/03/2024

WorkList ID : 176914

Department : Wet-Chemistry

Date : 01-03-2024 08:29:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1002-12	1155	Solid	Flash Point	Cool 4 deg C	PSEG03	O11	01/02/2024	1010B

Date/Time 01/04/2024 (8:30)
Raw Sample Received by: N.P. WC
Raw Sample Relinquished by: CR SR

Date/Time 01/04/2024 10:45
Raw Sample Received by: CR SR
Raw Sample Relinquished by: N.P.W.C