



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129037

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/04/2024	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	01/04/2024	17:11
3	CAL3	1	Water	NA	NA	20.3	10.02	01/04/2024	17:12
4	ICV	1	Water	NA	NA	20.5	6.98	01/04/2024	17:15
5	CCV1	1	Water	NA	NA	20.3	2.02	01/04/2024	17:46
6	P1015-02	1	Solid	20.02	20	23.3	9.63	01/04/2024	17:50
7	P1015-02DUP	1	Solid	20.03	20	23.4	9.64	01/04/2024	17:51
8	P1015-04	1	Solid	20.05	20	23.6	9.50	01/04/2024	17:55
9	P1029-01	1	Solid	20.02	20	22.7	6.66	01/04/2024	18:00
10	P1030-01	1	Solid	20.01	20	22.7	6.65	01/04/2024	18:05
11	P1030-02	1	Solid	20.02	20	22.3	7.72	01/04/2024	18:10
12	P1032-01	1	Solid	20.01	20	20.3	4.86	01/04/2024	18:15
13	P1032-02	1	Solid	20.02	20	22.6	8.27	01/04/2024	18:20
14	P1032-03	1	Solid	20.03	20	23.2	7.62	01/04/2024	18:25
15	CCV2	1	Water	NA	NA	20.3	12.02	01/04/2024	18:30

129037

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion p1021 WorkList ID : 176961 Department : Wet-Chemistry Date : 01-04-2024 11:07:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1015-02	MR-301P-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	01/02/2024	9045D
P1015-04	MR-301P-25-27	Solid	Corrosivity	Cool 4 deg C	HDR108	O11	01/03/2024	9045D
P1029-01	225	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	01/04/2024	9045D
P1030-01	3366-3370	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	01/04/2024	9045D
P1030-02	3374	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	01/04/2024	9045D
P1032-01	378	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	01/04/2024	9045D
P1032-02	379	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	01/04/2024	9045D
P1032-03	380	Solid	Corrosivity	Cool 4 deg C	PSEG03	O11	01/04/2024	9045D

Date/Time 01/04/24 17:00
 Raw Sample Received by: 20 (690)
 Raw Sample Relinquished by: 20 (690)

Date/Time 01/04/24 19:30
 Raw Sample Received by: 20 (690)
 Raw Sample Relinquished by: 20 (690)