

LB129040


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40123



A= 0.0000e+000

B= 2.4182e-004

C= -5.8732e-002

Rho= 0.9980565

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.022	-0.022	151	0.000	151	0				1
0.2	0.200	0.194	-0.006	1045	0.0 %	1045					3
2.5	2.500	2.357	-0.143	9990	0.0 %	9990					6
5.0	5.000	5.141	0.141	21502	0.0 %	21502					3
7.5	7.500	7.887	0.387	32857	0.0 %	32857					5
10.0	10.000	9.644	-0.356	40123	0.0 %	40123					4

LB129040 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jan 2024 08:03:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	151	-	0.0000	7470A	PPB	05 Jan 2024 08:07:37	S	Std
0.2 - 1	502	1045	-	0.2000	7470A	PPB	05 Jan 2024 08:09:55	S	Std
2.5 - 1	525	9990	-	2.5000	7470A	PPB	05 Jan 2024 08:12:11	S	Std
5.0 - 1	55	21502	-	5.0000	7470A	PPB	05 Jan 2024 08:14:29	S	Std
7.5 - 1	57.5	32857	-	7.5000	7470A	PPB	05 Jan 2024 08:16:46	S	Std
10.0 - 1	510	40123	-	10.0000	7470A	PPB	05 Jan 2024 08:19:05	S	Std
ICV52 - 1	ICV52	16846	4.0149	-	7470A	PPB	05 Jan 2024 08:27:02	U	SMPL
ICB52 - 1	ICB52	-123	-0.0885	-	7470A	PPB	05 Jan 2024 08:29:18	U	SMPL
CCV25 - 1	CCV25	21217	5.0719	-	7470A	PPB	05 Jan 2024 08:31:36	U	SMPL
CCB25 - 1	CCB25	-146	-0.0940	-	7470A	PPB	05 Jan 2024 08:33:51	U	SMPL
CRA - 1	CRA	968	0.1753	-	7470A	PPB	05 Jan 2024 08:36:09	U	SMPL
HighStd - 1	HighStd	42009	10.0998	-	7470A	PPB	05 Jan 2024 08:39:53	U	SMPL
ChkStd - 1	ChkStd	31025	7.4437	-	7470A	PPB	05 Jan 2024 08:42:07	U	SMPL
PB158257BL - 1	PBW	227	-0.0038	-	7470A	PPB	05 Jan 2024 08:47:06	U	SMPL
PB158257BS - 1	LCSW	17611	4.1999	-	7470A	PPB	05 Jan 2024 08:49:21	U	SMPL
P1003-02 - 1	72-11978	-61	-0.0735	-	7470A	PPB	05 Jan 2024 08:51:37	U	SMPL
P1003-02DUP - 1	72-11978DUP	124	-0.0287	-	7470A	PPB	05 Jan 2024 08:53:56	U	SMPL
P1003-02MS - 1	72-11978MS	18789	4.4848	-	7470A	PPB	05 Jan 2024 08:56:14	U	SMPL
P1003-02MSD - 1	72-11978MSD	18169	4.3349	-	7470A	PPB	05 Jan 2024 08:58:31	U	SMPL
PB158258BL - 1	PBW	123	-0.0290	-	7470A	PPB	05 Jan 2024 09:00:49	U	SMPL
CCV26 - 1	CCV26	21548	5.1520	-	7470A	PPB	05 Jan 2024 09:03:07	U	SMPL
CCB26 - 1	CCB26	-146	-0.0940	-	7470A	PPB	05 Jan 2024 09:05:22	U	SMPL
PB158258BS - 1	LCSW	19503	4.6574	-	7470A	PPB	05 Jan 2024 09:07:39	U	SMPL
O6038-07 - 1	EQ-TANK	8	-0.0568	-	7470A	PPB	05 Jan 2024 09:09:55	U	SMPL
P1014-02 - 1	COMP	586	0.0830	-	7470A	PPB	05 Jan 2024 09:12:13	U	SMPL
P1014-02DUP - 1	COMPDUP	667	0.1026	-	7470A	PPB	05 Jan 2024 09:14:28	U	SMPL
P1014-02MS - 1	COMPMS	12042	2.8532	-	7470A	PPB	05 Jan 2024 09:16:44	U	SMPL
P1014-02MSD - 1	COMPMSD	12344	2.9263	-	7470A	PPB	05 Jan 2024 09:19:02	U	SMPL
CCV27 - 1	CCV27	19498	4.6562	-	7470A	PPB	05 Jan 2024 09:23:41	U	SMPL
CCB27 - 1	CCB27	-165	-0.0986	-	7470A	PPB	05 Jan 2024 09:25:58	U	SMPL
P1003-02LX5 - 1		26	-0.0524	-	7470A	PPB	05 Jan 2024 09:28:16	U	SMPL
P1003-02A - 1		18525	4.4209	-	7470A	PPB	05 Jan 2024 09:30:32	U	SMPL
P1014-02LX5 - 1		-12	-0.0616	-	7470A	PPB	05 Jan 2024 09:32:47	U	SMPL
P1014-02A - 1		18517	4.4190	-	7470A	PPB	05 Jan 2024 09:35:06	U	SMPL
PB158198TB - 1	PB158198TB	355	0.0271	-	7470A	PPB	05 Jan 2024 11:47:47	U	SMPL
CCV28 - 1	CCV28	22569	5.3989	-	7470A	PPB	05 Jan 2024 11:50:22	U	SMPL
CCB28 - 1	CCB28	178	-0.0157	-	7470A	PPB	05 Jan 2024 11:55:21	U	SMPL