



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129041

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	01/05/2024	10:30
2	P1028-01		1	12	100.00	>212.0	>212.0	01/05/2024	11:00
3	P1028-01DUP		1	14	100.00	>212.0	>212.0	01/05/2024	11:30
4	P1028-02		1	14	100.00	>212.0	>212.0	01/05/2024	12:00
5	P1029-02		1	12	100.00	>212.0	>212.0	01/05/2024	12:30
6	P1029-03		1	12	100.00	>212.0	>212.0	01/05/2024	13:00
7	P1030-03		1	14	100.00	>212.0	>212.0	01/05/2024	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129041

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint 01/05/2024

WorkList ID : 176980

Department : Wet-Chemistry

Date : 01-05-2024 08:03:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1028-01	2948-2950	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/04/2024	1010B
P1028-02	2956	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/04/2024	1010B
P1029-02	228	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/04/2024	1010B
P1029-03	229	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/04/2024	1010B
P1030-03	3253-3375	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/04/2024	1010B

Date/Time 01/05/2024 (10:00)

Raw Sample Received by: N.P. WC

Raw Sample Relinquished by: N.P. WC

Date/Time 01/05/2024 13:45

Raw Sample Received by: N.P. WC

Raw Sample Relinquished by: N.P. WC