



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129043

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/05/2024	09:20
2	CAL2	1	Water	NA	NA	20.3	7.00	01/05/2024	09:21
3	CAL3	1	Water	NA	NA	20.3	10.02	01/05/2024	09:22
4	ICV	1	Water	NA	NA	20.5	7.01	01/05/2024	09:25
5	CCV1	1	Water	NA	NA	20.3	2.01	01/05/2024	09:30
6	P1026-01	1	Water	NA	NA	20.6	4.73	01/05/2024	09:35
7	P1026-01DUP	1	Water	NA	NA	20.7	4.74	01/05/2024	09:36
8	CCV2	1	Water	NA	NA	20.5	12.03	01/05/2024	09:38

WORKLIST(Hardcopy Internal Chain)

129013

WorkList Name : PH P1026

WorkList ID : 176990

Department : Wet-Chemistry

Date : 01-05-2024 09:10:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1026-01	☒ HAW-RIVER-MAIN-CELL	Water	pH	Cool 4 deg C	STER11	O11	01/02/2024	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: