

LB129045


7471B

INSTRUMENT ID: CN1

Linear

 μ Abs.:

58521



A= 0.0000e+000

B= 1.7128e-004

C= -2.5629e-003

Rho= 0.9999689

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ (s.d)
0.0	0.000	0.009	0.009	66	0.000	66					2
0.2	0.200	0.205	0.005	1212	0.0 %	1212					1
2.5	2.500	2.523	0.023	14743	0.0 %	14743					0
5.0	5.000	4.937	-0.063	28838	0.0 %	28838					0
7.5	7.500	7.506	0.006	43839	0.0 %	43839					0
10.0	10.000	10.021	0.021	58521	0.0 %	58521					0

LB129045 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jan 2024 13:13:39

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	66	-	0.00007471B	PPB		05 Jan 2024 13:15:24	S	Std
0.2 - 1	50-2	1212	-	0.20007471B	PPB		05 Jan 2024 13:17:41	S	Std
2.5 - 1	52-5	14743	-	2.50007471B	PPB		05 Jan 2024 13:19:58	S	Std
5.0 - 1	55	28838	-	5.00007471B	PPB		05 Jan 2024 13:22:15	S	Std
7.5 - 1	57-5	43839	-	7.50007471B	PPB		05 Jan 2024 13:24:34	S	Std
10.0 - 1	5N	58521	-	10.00007471B	PPB		05 Jan 2024 13:27:01	S	Std
ICV53 - 1	ICV53	23418	4.0084	-7471B	PPB		05 Jan 2024 13:33:31	U	SMPL
ICB53 - 1	ICB53	-162	-0.0303	-7471B	PPB		05 Jan 2024 13:35:45	U	SMPL
CCV29 - 1	CCV29	29721	5.0880	-7471B	PPB		05 Jan 2024 13:38:08	U	SMPL
CCB29 - 1	CCB29	-171	-0.0319	-7471B	PPB		05 Jan 2024 13:40:24	U	SMPL
CRA - 1	CRA	1282	0.2170	-7471B	PPB		05 Jan 2024 13:42:50	U	SMPL
HighStd - 1	HighStd	60172	10.3036	-7471B	PPB		05 Jan 2024 13:45:06	U	SMPL
ChkStd - 1	ChkStd	41887	7.1718	-7471B	PPB		05 Jan 2024 13:47:21	U	SMPL
PB158279BL - 1	PBS	-211	-0.0387	-7471B	PPB		05 Jan 2024 13:49:53	U	SMPL
PB158279BS - 1	LCSS	23454	4.0146	-7471B	PPB		05 Jan 2024 13:55:04	U	SMPL
P1002-11 - 1	1132	-83	-0.0168	-7471B	PPB		05 Jan 2024 13:57:21	U	SMPL
P1003-01 - 1	72-11978	24327	4.1641	-7471B	PPB		05 Jan 2024 13:59:42	U	SMPL
P1004-01 - 1	SU-04-01022024	1116	0.1886	-7471B	PPB		05 Jan 2024 14:01:59	U	SMPL
P1005-01 - 1	RBR200030	3186	0.5431	-7471B	PPB		05 Jan 2024 14:04:24	U	SMPL
P1006-01 - 1	NB-7-010324	957	0.1614	-7471B	PPB		05 Jan 2024 14:06:40	U	SMPL
CCV30 - 1	CCV30	30121	5.1565	-7471B	PPB		05 Jan 2024 14:08:56	U	SMPL
CCB30 - 1	CCB30	-168	-0.0313	-7471B	PPB		05 Jan 2024 14:11:13	U	SMPL
P1006-01DUP - 1	NB-7-010324DUP	1092	0.1845	-7471B	PPB		05 Jan 2024 14:16:36	U	SMPL
P1006-01MS - 1	NB-7-010324MS	23233	3.9767	-7471B	PPB		05 Jan 2024 14:22:41	U	SMPL
P1006-01MSD ; 1	NB-7-010324MSD	23069	3.9486	-7471B	PPB		05 Jan 2024 14:24:56	U	SMPL
P1015-01 - 1	MR-301P-0-2	9276	1.5862	-7471B	PPB		05 Jan 2024 14:27:19	U	SMPL
P1015-03 - 1	MR-301P-25-27	1019	0.1720	-7471B	PPB		05 Jan 2024 14:29:44	U	SMPL
P1040-01 - 1	COMP-1	78987	13.5262	-7471B	PPB		05 Jan 2024 14:32:02	U	SMPL
P1040-02 - 1	COMP-2	3341	0.5697	-7471B	PPB		05 Jan 2024 14:34:19	U	SMPL
P1040-03 - 1	COMP-3	823	0.1384	-7471B	PPB		05 Jan 2024 14:36:53	U	SMPL
P1006-01LX5 - 1		121	0.0182	-7471B	PPB		05 Jan 2024 14:39:12	U	SMPL
P1006-01A - 1		23660	4.0499	-7471B	PPB		05 Jan 2024 14:41:28	U	SMPL
CCV31 - 1	CCV31	30381	5.2010	-7471B	PPB		05 Jan 2024 14:43:43	U	SMPL
CCB31 - 1	CCB31	-180	-0.0334	-7471B	PPB		05 Jan 2024 14:46:08	U	SMPL
P1040-01DLX5 - 1		17339	2.9672	-7471B	PPB		05 Jan 2024 14:48:32	U	SMPL
CCV32 - 1	CCV32	30528	5.2262	-7471B	PPB		05 Jan 2024 14:50:48	U	SMPL
CCB32 - 1	CCB32	-199	-0.0366	-7471B	PPB		05 Jan 2024 14:53:06	U	SMPL