



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB129050

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	01/05/2024	16:15
2	CAL2	1	Water	NA	NA	20.4	6.99	01/05/2024	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	01/05/2024	16:19
4	ICV	1	Water	NA	NA	20.3	7.01	01/05/2024	16:20
5	CCV1	1	Water	NA	NA	20.2	2.01	01/05/2024	16:24
6	P1049-01	1	Water	NA	NA	20.7	6.28	01/05/2024	16:30
7	P1049-03	1	Water	NA	NA	20.7	5.59	01/05/2024	16:35
8	P1049-04	1	Water	NA	NA	20.4	5.62	01/05/2024	16:38
9	P1049-06	1	Water	NA	NA	20.2	6.60	01/05/2024	16:45
10	P1049-07	1	Water	NA	NA	20.1	6.51	01/05/2024	16:55
11	P1049-09	1	Water	NA	NA	20.2	5.72	01/05/2024	17:00
12	P1049-10	1	Water	NA	NA	20.3	5.61	01/05/2024	17:05
13	P1049-11	1	Water	NA	NA	20.3	5.82	01/05/2024	17:15
14	P1049-11DUP	1	Water	NA	NA	20.4	5.83	01/05/2024	17:16
15	CCV2	1	Water	NA	NA	20.3	12.02	01/05/2024	17:20

129050

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 1049 WorkList ID : 177001 Department : Wet-Chemistry Date : 01-05-2024 16:02:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1049-01	50249	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C
P1049-03	50220	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C
P1049-04	50221	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C
P1049-06	50250	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C
P1049-07	50251	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C
P1049-09	50256	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C
P1049-10	50257	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C
P1049-11	50258	Water	pH	Cool 4 deg C	PSEG03	O31	01/05/2024	9040C

Date/Time 01/05/24 96:110
Raw Sample Received by: RD LWC1
Raw Sample Relinquished by: GR CARM

Date/Time 01/05/24 17:30
Raw Sample Received by: GR CARM
Raw Sample Relinquished by: RD LWC1