

LB129055


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

5245E



A= 0.0000e+000

B= 1.8801e-004 slope

C= -1.5515e-002 y-intercept

Rho= 0.9995530

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	of 10
0.00	0.000	-0.002	-0.002	74	0.000	74					=
0.05	0.050					442	0				-5
0.20	0.200	0.191	-0.009	1098	0.0 %	1098					-2
2.50	2.500	2.462	-0.038	13177	0.0 %	13177					0
5.00	5.000	4.984	-0.016	26589	0.0 %	26589					3
7.50	7.500	7.718	0.218	41136	0.0 %	41136					-2
10.0	10.000	9.847	-0.153	52456	0.0 %	52456					

LB129055

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	74	-		0 SFAM01.1	PPB	1/8/2024 10:03	Std
	0.2 S01	1098	-		0.2 SFAM01.1	PPB	1/8/2024 10:06	Std
	2.5 S02	13177	-		2.5 SFAM01.1	PPB	1/8/2024 10:10	Std
	5 S03	26589	-		5 SFAM01.1	PPB	1/8/2024 10:12	Std
	7.5 S04	41136	-		7.5 SFAM01.1	PPB	1/8/2024 10:14	Std
	10 S05	52456	-		10 SFAM01.1	PPB	1/8/2024 10:19	Std
ICV054	ICV054	21673	4.0592 -		SFAM01.1	PPB	1/8/2024 10:22	SMPL
ICB054	ICB054	-205	-0.0541 -		SFAM01.1	PPB	1/8/2024 10:24	SMPL
CCV033	CCV033	27664	5.1856 -		SFAM01.1	PPB	1/8/2024 10:27	SMPL
CCB033	CCB033	-216	-0.0561 -		SFAM01.1	PPB	1/8/2024 10:29	SMPL
PB158296BL	PBS296	67	-0.0029 -		SFAM01.1	PPB	1/8/2024 10:34	SMPL
O5893-01	E0XD7	5621	1.0413 -		SFAM01.1	PPB	1/8/2024 10:36	SMPL
O5893-02	E0XD8	1241	0.2178 -		SFAM01.1	PPB	1/8/2024 10:38	SMPL
O5893-03	E0XF9	2409	0.4374 -		SFAM01.1	PPB	1/8/2024 10:40	SMPL
O5893-04	E0XG0	1799	0.3227 -		SFAM01.1	PPB	1/8/2024 10:43	SMPL
O5893-05	E0XG2	934	0.1601 -		SFAM01.1	PPB	1/8/2024 10:45	SMPL
O5893-06	E0XX0	874	0.1488 -		SFAM01.1	PPB	1/8/2024 10:47	SMPL
O5893-07	E0XX1	822	0.139 -		SFAM01.1	PPB	1/8/2024 10:49	SMPL
O5893-08	E0XH1	570	0.0917 -		SFAM01.1	PPB	1/8/2024 10:52	SMPL
O5893-09	E0XH2	638	0.1044 -		SFAM01.1	PPB	1/8/2024 10:54	SMPL
O5893-10	E0XH3	283	0.0377 -		SFAM01.1	PPB	1/8/2024 10:56	SMPL
O5893-11	E0XH3D	331	0.0467 -		SFAM01.1	PPB	1/8/2024 10:59	SMPL
O5893-13	E0XH9	483	0.0753 -		SFAM01.1	PPB	1/8/2024 11:03	SMPL
O5893-14	E0XJ0	3695	0.6792 -		SFAM01.1	PPB	1/8/2024 11:05	SMPL
O5893-15	E0XJ3	478	0.0744 -		SFAM01.1	PPB	1/8/2024 11:08	SMPL
O5893-16	E0X99	270	0.0352 -		SFAM01.1	PPB	1/8/2024 11:10	SMPL
O5893-17	E0XA0	506	0.0796 -		SFAM01.1	PPB	1/8/2024 11:12	SMPL
O5893-18	E0XA1	363	0.0527 -		SFAM01.1	PPB	1/8/2024 11:14	SMPL
O5893-19	E0XA2	309	0.0426 -		SFAM01.1	PPB	1/8/2024 11:17	SMPL
CCV034	CCV034	27059	5.0719 -		SFAM01.1	PPB	1/8/2024 11:19	SMPL
CCB034	CCB034	-153	-0.0443 -		SFAM01.1	PPB	1/8/2024 11:21	SMPL
O5893-20	E0XJ1	228	0.0274 -		SFAM01.1	PPB	1/8/2024 11:24	SMPL
O5893-21	E0XJ2	254	0.0322 -		SFAM01.1	PPB	1/8/2024 11:26	SMPL
O5893-22	E0X53	1207	0.2114 -		SFAM01.1	PPB	1/8/2024 11:28	SMPL
PB158297BL	PBS297	-32	-0.0215 -		SFAM01.1	PPB	1/8/2024 11:30	SMPL
O5922-01	E0X54	1745	0.3126 -		SFAM01.1	PPB	1/8/2024 11:33	SMPL
O5922-02	E0XJ4	292	0.0394 -		SFAM01.1	PPB	1/8/2024 11:35	SMPL
O5922-03	E0XJ5	502	0.0789 -		SFAM01.1	PPB	1/8/2024 11:37	SMPL
O5922-07	E0X51	840	0.1424 -		SFAM01.1	PPB	1/8/2024 11:47	SMPL
O5922-08	E0X83	508	0.08 -		SFAM01.1	PPB	1/8/2024 11:49	SMPL
O5922-09	E0XK1	572	0.092 -		SFAM01.1	PPB	1/8/2024 11:52	SMPL
O5922-10	E0XK2	663	0.1091 -		SFAM01.1	PPB	1/8/2024 11:54	SMPL
O5922-11	E0XK3	646	0.1059 -		SFAM01.1	PPB	1/8/2024 11:56	SMPL
O5922-12	E0XK4	7034	1.307 -		SFAM01.1	PPB	1/8/2024 11:58	SMPL
O5922-13	E0XX7	124	0.0078 -		SFAM01.1	PPB	1/8/2024 12:01	SMPL
O5922-14	E0X52	508	0.08 -		SFAM01.1	PPB	1/8/2024 12:03	SMPL

LB129055

INSTRUMENT ID : CV1

O5922-16	E0X84	5375	0.995 -	SFAM01.1	PPB	1/8/2024 12:05 SMPL
O5922-17	E0X85	152	0.0131 -	SFAM01.1	PPB	1/8/2024 12:07 SMPL
CCV035	CCV035	26542	4.9747 -	SFAM01.1	PPB	1/8/2024 12:10 SMPL
CCB035	CCB035	-103	-0.0349 -	SFAM01.1	PPB	1/8/2024 12:12 SMPL
O5922-18	E0X87	919	0.1573 -	SFAM01.1	PPB	1/8/2024 12:14 SMPL
O5922-19	E0X88	318	0.0443 -	SFAM01.1	PPB	1/8/2024 12:17 SMPL
O5922-21	E0X86	421	0.0636 -	SFAM01.1	PPB	1/8/2024 12:19 SMPL
O5922-22	E0X89	253	0.0321 -	SFAM01.1	PPB	1/8/2024 12:21 SMPL
O5893-12	E0XH3S	13182	2.4628 -	SFAM01.1	PPB	1/8/2024 12:23 SMPL
O5922-04	E0XX3	646	0.1059 -	SFAM01.1	PPB	1/8/2024 12:26 SMPL
O5922-05	E0XX3D	691	0.1144 -	SFAM01.1	PPB	1/8/2024 12:28 SMPL
O5922-06	E0XX3S	11570	2.1598 -	SFAM01.1	PPB	1/8/2024 12:30 SMPL
CCV036	CCV036	26823	5.0275 -	SFAM01.1	PPB	1/8/2024 12:32 SMPL
CCB036	CCB036	-90	-0.0324 -	SFAM01.1	PPB	1/8/2024 12:35 SMPL