


LB129061

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 1.9651e-004 slope
 C= -1.1037e-002 intercept
 Rho= 0.9999348
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	CV
0.00	0.000	-0.005	-0.005	32	0.000	32					11
0.05	0.050					442	0				8
0.20	0.200	0.184	-0.016	995	0.0 %	995					1
2.50	2.500	2.472	-0.028	12634	0.0 %	12634					2
5.00	5.000	5.092	0.092	25971	0.0 %	25971					0
7.50	7.500	7.484	-0.016	38142	0.0 %	38142					0
10.0	10.000	9.972	-0.028	50802	0.0 %	50802					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	32	-		0 SFAM01.1	PPB	1/8/2024 13:14	Std
	0.2 S01	995	-		0.2 SFAM01.1	PPB	1/8/2024 13:16	Std
	2.5 S02	12634	-		2.5 SFAM01.1	PPB	1/8/2024 13:20	Std
	5 S03	25971	-		5 SFAM01.1	PPB	1/8/2024 13:23	Std
	7.5 S04	38142	-		7.5 SFAM01.1	PPB	1/8/2024 13:25	Std
	10 S05	50802	-		10 SFAM01.1	PPB	1/8/2024 13:27	Std
ICV055	ICV055	19714	3.8629 -		SFAM01.1	PPB	1/8/2024 13:30	SMPL
ICB055	ICB055	-138	-0.0382 -		SFAM01.1	PPB	1/8/2024 13:32	SMPL
CCV037	CCV037	26141	5.1259 -		SFAM01.1	PPB	1/8/2024 13:35	SMPL
CCB037	CCB037	-157	-0.0419 -		SFAM01.1	PPB	1/8/2024 13:37	SMPL
PB158298BL	PBW298	20	-0.0071 -		SFAM01.1	PPB	1/8/2024 13:39	SMPL
O5871-01	E0XQ4	17	-0.0077 -		SFAM01.1	PPB	1/8/2024 13:41	SMPL
O5871-02	E0XQ5	23	-0.0065 -		SFAM01.1	PPB	1/8/2024 13:44	SMPL
O5871-03	E0XR4	21	-0.0069 -		SFAM01.1	PPB	1/8/2024 13:46	SMPL
O5871-04	E0XR5	27	-0.0057 -		SFAM01.1	PPB	1/8/2024 13:48	SMPL
O5871-05	E0XQ6	14	-0.0083 -		SFAM01.1	PPB	1/8/2024 13:50	SMPL
O5871-06	E0XR6	32	-0.0047 -		SFAM01.1	PPB	1/8/2024 13:53	SMPL
O5871-07	E0XQ7	10	-0.0091 -		SFAM01.1	PPB	1/8/2024 13:55	SMPL
O5871-08	E0XR7	31	-0.0049 -		SFAM01.1	PPB	1/8/2024 13:57	SMPL
O5871-14	E0XQ8	28	-0.0055 -		SFAM01.1	PPB	1/8/2024 14:00	SMPL
O5871-15	E0XR8	20	-0.0071 -		SFAM01.1	PPB	1/8/2024 14:02	SMPL
O5871-16	E0X22	35	-0.0042 -		SFAM01.1	PPB	1/8/2024 14:04	SMPL
O5871-17	E0XX4	29	-0.0053 -		SFAM01.1	PPB	1/8/2024 14:06	SMPL
O5871-18	E0XX5	21	-0.0069 -		SFAM01.1	PPB	1/8/2024 14:09	SMPL
PB158299BL	PBW299	49	-0.0014 -		SFAM01.1	PPB	1/8/2024 14:11	SMPL
O6001-01	MC0AF5	62	0.0011 -		SFAM01.1	PPB	1/8/2024 14:13	SMPL
O6001-02	MC0AF7	32	-0.0047 -		SFAM01.1	PPB	1/8/2024 14:16	SMPL
O6001-03	MC0AF9	27	-0.0057 -		SFAM01.1	PPB	1/8/2024 14:18	SMPL
O6001-04	MC0AG0	157	0.0198 -		SFAM01.1	PPB	1/8/2024 14:20	SMPL
O6001-05	MC0AG1	481	0.0835 -		SFAM01.1	PPB	1/8/2024 14:22	SMPL
CCV038	CCV038	25574	5.0145 -		SFAM01.1	PPB	1/8/2024 14:25	SMPL
CCB038	CCB038	-149	-0.0403 -		SFAM01.1	PPB	1/8/2024 14:27	SMPL
O6001-06	MC0AG1D	-2	-0.0114 -		SFAM01.1	PPB	1/8/2024 14:29	SMPL
O6001-07	MC0AG1S	6086	1.1849 -		SFAM01.1	PPB	1/8/2024 14:34	SMPL
O6001-08	MC0AG2	-9	-0.0128 -		SFAM01.1	PPB	1/8/2024 14:36	SMPL
O6001-09	MC0AG3	9	-0.0093 -		SFAM01.1	PPB	1/8/2024 14:39	SMPL
O6001-10	MC0AG4	5	-0.0101 -		SFAM01.1	PPB	1/8/2024 14:41	SMPL
O6001-11	MC0AK7	-10	-0.013 -		SFAM01.1	PPB	1/8/2024 14:43	SMPL
O6001-12	MC0AM7	3	-0.0104 -		SFAM01.1	PPB	1/8/2024 14:45	SMPL
CCV039	CCV039	25394	4.9791 -		SFAM01.1	PPB	1/8/2024 14:48	SMPL
CCB039	CCB039	-168	-0.0441 -		SFAM01.1	PPB	1/8/2024 14:50	SMPL