

LB129069


245.1

INSTRUMENT ID: EVJ

Linear

 μ Abs.:

42418



A= 0.0000e+000

B= 2.3908e-004

C= -2.0609e-001

Rho= 0.9998708

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	b/o
0.0	0.000	0.017	0.017	932	0.000	932					0
0.05	0.050	0.059	0.009	1107	0.0 %	1107	0				18
0.2	0.200	0.213	0.013	1753	0.0 %	1753					6
2.5	2.500	2.399	-0.101	10895	0.0 %	10895					-4
5.0	5.000	5.034	0.034	21918	0.0 %	21918					1
7.5	7.500	7.594	0.094	32624	0.0 %	32624					1
10.0	10.000	9.935	-0.065	42418	0.0 %	42418					-1

LB129069 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Jan 2024 09:48:03

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.9.05	932	-	0.0000	245.1	PPB	09 Jan 2024 10:12:18	S	Std
0.05 - 1	50.9.05	1107	-	0.0500	245.1	PPB	09 Jan 2024 10:14:36	S	Std
0.2 - 1	52.5	1753	-	0.2000	245.1	PPB	09 Jan 2024 10:16:53	S	Std
2.5 - 1	55	10895	-	2.5000	245.1	PPB	09 Jan 2024 10:19:11	S	Std
5.0 - 1	55	21918	-	5.0000	245.1	PPB	09 Jan 2024 10:21:28	S	Std
7.5 - 1	57.5	32624	-	7.5000	245.1	PPB	09 Jan 2024 10:23:46	S	Std
10.0 - 1	51.1	42418	-	10.0000	245.1	PPB	09 Jan 2024 10:31:26	S	Std
ICV56 - 1	ICV56	17319	3.9345	-245.1		PPB	09 Jan 2024 10:40:08	U	SMPL
ICB56 - 1	ICB56	657	-0.0490	-245.1		PPB	09 Jan 2024 10:44:59	U	SMPL
CCV40 - 1	CCV40	22077	5.0721	-245.1		PPB	09 Jan 2024 10:47:14	U	SMPL
CCB40 - 1	CCB40	826	-0.0086	-245.1		PPB	09 Jan 2024 10:49:29	U	SMPL
CRA - 1	CRA	1066	0.0486	-245.1		PPB	09 Jan 2024 10:54:35	U	SMPL
HighStd - 1	HighStd	42738	10.0117	-245.1		PPB	09 Jan 2024 10:56:49	U	SMPL
ChkStd - 1	ChkStd	29721	6.8996	-245.1		PPB	09 Jan 2024 10:59:04	U	SMPL
PB158318BL - 1	PBW	693	-0.0404	-245.1		PPB	09 Jan 2024 11:01:28	U	SMPL
PB158318BS - 1	LCSW	16182	3.6627	-245.1		PPB	09 Jan 2024 11:05:55	U	SMPL
O4699-10 - 1	LOD-MDL-WATER-04-QT4-2023	1057	0.0486	-245.1		PPB	09 Jan 2024 11:10:47	U	SMPL
O4699-11 - 1	LOQ-WATER-05-QT4-2023	1691	0.1982	-245.1		PPB	09 Jan 2024 11:13:03	U	SMPL
P1008-01 - 1	402	0	-0.2061	-245.1		PPB	09 Jan 2024 11:17:56	U	SMPL
P1008-01DUP - 1	402DUP	3	-0.2054	-245.1		PPB	09 Jan 2024 11:20:12	U	SMPL
P1008-01MS - 1	402MS	16422	3.7201	-245.1		PPB	09 Jan 2024 11:24:27	U	SMPL
CCV41 - 1	CCV41	22683	5.2170	-245.1		PPB	09 Jan 2024 11:26:42	U	SMPL
CCB41 - 1	CCB41	825	-0.0089	-245.1		PPB	09 Jan 2024 11:29:00	U	SMPL
P1008-01MSD - 1	402MSD	16283	3.6869	-245.1		PPB	09 Jan 2024 11:31:18	U	SMPL
P1008-01LX5 - 1		319	-0.1298	-245.1		PPB	09 Jan 2024 11:33:38	U	SMPL
P1008-01A - 1		19124	4.3661	-245.1		PPB	09 Jan 2024 11:35:56	U	SMPL
CCV42 - 1	CCV42	22460	5.1637	-245.1		PPB	09 Jan 2024 11:38:12	U	SMPL
CCB42 - 1	CCB42	799	-0.0151	-245.1		PPB	09 Jan 2024 11:40:30	U	SMPL