



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129074

Slope : 99.2

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3070         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W2981         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3043         |
| buffer solution pH 7 yellow    | W3071         |
| Buffer Solution, PH2 (500ml)   | W3005         |
| Buffer Solution, PH12 (500ml)  | W3072         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.3             | 4.01        | 01/09/2024 | 11:35     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.3             | 6.99        | 01/09/2024 | 11:36     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.02       | 01/09/2024 | 11:38     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.3             | 7.02        | 01/09/2024 | 11:40     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.3             | 2.02        | 01/09/2024 | 11:45     |
| 6   | P1076-01    | 1  | Water  | NA          | NA          | 20.1             | 5.95        | 01/09/2024 | 11:50     |
| 7   | P1076-01DUP | 1  | Water  | NA          | NA          | 20.2             | 5.96        | 01/09/2024 | 11:51     |
| 8   | P1076-02    | 1  | Water  | NA          | NA          | 20.4             | 6.01        | 01/09/2024 | 11:55     |
| 9   | CCV2        | 1  | Water  | NA          | NA          | 20.3             | 12.02       | 01/09/2024 | 12:00     |

# WORKLIST(Hardcopy Internal Chain)

W 129074

WorkList Name : ph o1076

WorkList ID : 177048

Department : Wet-Chemistry

Date : 01-09-2024 11:26:31

| Sample   | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|------|--------------|----------|-----------------------------|--------------|--------|
| P1076-01 | A MAIN-CELL     | Water  | pH   | Cool 4 deg C | STER16   | O11                         | 01/05/2024   | 9040C  |
| P1076-02 | A EX-CELL       | Water  | pH   | Cool 4 deg C | STER16   | O11                         | 01/05/2024   | 9040C  |

Date/Time 01/09/24 11:30

Raw Sample Received by: SD (COC)

Raw Sample Relinquished by: JDCSM

Date/Time 01/09/24

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: JDCSM

13:30

JDCSM