



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129075

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/09/2024	11:35
2	CAL2	1	Water	NA	NA	20.3	6.99	01/09/2024	11:36
3	CAL3	1	Water	NA	NA	20.3	10.02	01/09/2024	11:38
4	ICV	1	Water	NA	NA	20.3	7.02	01/09/2024	11:40
5	CCV1	1	Water	NA	NA	20.3	2.02	01/09/2024	12:05
6	P1067-02	1	Solid	20.02	20	22.5	5.85	01/09/2024	12:10
7	P1067-05	1	Solid	20.03	20	22.9	7.91	01/09/2024	12:17
8	P1067-08	1	Solid	20.02	20	21.5	7.47	01/09/2024	12:25
9	P1067-08DUP	1	Solid	20.03	20	21.6	7.48	01/09/2024	12:26
10	CCV2	1	Water	NA	NA	20.3	12.02	01/09/2024	12:30

129045

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion p1077 WorkList ID : 177046 Department : Wet-Chemistry Date : 01-09-2024 11:17:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1067-02	SP-A-COMP(0-8)	Solid	Corrosivity	Cool 4 deg C	EART12	O11	01/08/2024	9045D
P1067-05	SP-B-COMP(0-8)	Solid	Corrosivity	Cool 4 deg C	EART12	O11	01/08/2024	9045D
P1067-08	SP-C-COMP(0-8)	Solid	Corrosivity	Cool 4 deg C	EART12	O11	01/08/2024	9045D

Date/Time 01/09/24 11:30
Raw Sample Received by: SD/60C
Raw Sample Relinquished by: SD CSM

Date/Time 01/09/24 11:00
Raw Sample Received by: SD CSM
Raw Sample Relinquished by: SD CSM