

LB 129077


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39692



A= 0.0000e+000

B= 2.4871e-004

C= -9.1245e-003

Rho= 0.9996771

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	a/b/d
0.0	0.000	-0.001	-0.001	33	0.000	33	0				-
0.2	0.200	0.188	-0.012	794	0.0 %	794					-6
2.5	2.500	2.431	-0.069	9810	0.0 %	9810					-3
5.0	5.000	5.059	0.059	20378	0.0 %	20378					1
7.5	7.500	7.660	0.160	30834	0.0 %	30834					2
10.0	10.000	9.863	-0.137	39692	0.0 %	39692					-1

LB129077 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Jan 2024 12:16:10

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	33	-	0.000	7470A	PPB	09 Jan 2024 13:12:46	S	Std
0.2 - 1	50.2	794	-	0.200	7470A	PPB	09 Jan 2024 13:15:02	S	Std
2.5 - 1	52.5	9810	-	2.500	7470A	PPB	09 Jan 2024 13:17:19	S	Std
5.0 - 1	55	20378	-	5.000	7470A	PPB	09 Jan 2024 13:19:36	S	Std
7.5 - 1	52.5	30834	-	7.500	7470A	PPB	09 Jan 2024 13:21:54	S	Std
10.0 - 1	510	39692	-	10.000	7470A	PPB	09 Jan 2024 13:32:13	S	Std
ICV57 - 1	ICV57	17128	4.2509	-	7470A	PPB	09 Jan 2024 13:34:56	U	SMPL
ICB57 - 1	ICB57	-299	-0.0835	-	7470A	PPB	09 Jan 2024 13:37:12	U	SMPL
CCV43 - 1	CCV43	20442	5.0751	-	7470A	PPB	09 Jan 2024 13:39:29	U	SMPL
CCB43 - 1	CCB43	-296	-0.0827	-	7470A	PPB	09 Jan 2024 13:41:46	U	SMPL
CRA - 1	CRA	750	0.1774	-	7470A	PPB	09 Jan 2024 13:44:03	U	SMPL
HighStd - 1	HighStd	40053	9.9526	-	7470A	PPB	09 Jan 2024 13:46:19	U	SMPL
ChkStd - 1	ChkStd	29513	7.3312	-	7470A	PPB	09 Jan 2024 13:48:34	U	SMPL
PB158302BL - 1	PBW	-70	-0.0265	-	7470A	PPB	09 Jan 2024 13:53:33	U	SMPL
PB158302BS - 1	LCSW	17302	4.2941	-	7470A	PPB	09 Jan 2024 13:58:10	U	SMPL
P1015-02 - 1	MR-301P-0-2	-33	-0.0173	-	7470A	PPB	09 Jan 2024 14:00:27	U	SMPL
P1015-02DUP - 1	MR-301P-0-2DUP	224	0.0466	-	7470A	PPB	09 Jan 2024 14:02:46	U	SMPL
P1015-02MS - 1	MR-301P-0-2MS	15410	3.8236	-	7470A	PPB	09 Jan 2024 14:07:36	U	SMPL
P1015-02MSD - 1	MR-301P-0-2MSD	15303	3.7970	-	7470A	PPB	09 Jan 2024 14:09:53	U	SMPL
P1015-04 - 1	MR-301P-25-27	-243	-0.0696	-	7470A	PPB	09 Jan 2024 14:12:10	U	SMPL
CCV44 - 1	CCV44	20091	4.9878	-	7470A	PPB	09 Jan 2024 14:14:28	U	SMPL
CCB44 - 1	CCB44	-307	-0.0855	-	7470A	PPB	09 Jan 2024 14:16:43	U	SMPL
P1039-01 - 1	TREMONT-AVE-DRUM-27-POLY	-66	-0.0255	-	7470A	PPB	09 Jan 2024 14:19:03	U	SMPL
P1039-02 - 1	TREMONT-AVE-DRUM-28-PPE	-39	-0.0188	-	7470A	PPB	09 Jan 2024 14:21:18	U	SMPL
PB158302TB - 1	PB158302TB	57	0.0051	-	7470A	PPB	09 Jan 2024 14:23:33	U	SMPL
P1015-02LX5 - 1		47	0.0026	-	7470A	PPB	09 Jan 2024 14:25:49	U	SMPL
P1015-02A - 1		16484	4.0907	-	7470A	PPB	09 Jan 2024 14:28:11	U	SMPL
CCV45 - 1	CCV45	20919	5.1937	-	7470A	PPB	09 Jan 2024 14:30:27	U	SMPL
CCB45 - 1	CCB45	-309	-0.0860	-	7470A	PPB	09 Jan 2024 14:32:46	U	SMPL