



Analytical Summary Report

Analysis Method: SM2120 B
Parameter: Color
Run Number: LB129078

Reviewed By: Rubina
Supervisor Review By: Iwona
pH Meter ID: WC pH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP106021
Calibration std, 5 color unit	WP106022
Calibration std, 10 color unit	WP106023
Calibration std, 15 color unit	WP106024
Calibration std, 20 color unit	WP106025
Calibration std, 25 color unit	WP106026
Calibration std, 30 color unit	WP106027
Calibration std, 40 color unit	WP106028
Calibration std, 50 color unit	WP106029
Calibration std, 60 color unit	WP106030
Calibration std, 70 color unit	WP106031
Calibration std, 100 color unit	WP106032

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB129078BL	LB129078BL	1	5.03	N/A	0.00	01/09/2024	13:00
2	P1078-09	7CS	20	7.49	N/A	60.00	01/09/2024	13:05
3	P1078-09DUP	7CSDUP	20	7.48	N/A	60.00	01/09/2024	13:10

WORKLIST(Hardcopy Internal Chain)

LB 123078

WorkList Name : COLOR-010923

WorkList ID : 177047

Department : Wet-Chemistry

Date : 01-09-2024 11:19:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1078-09	7CS	Water	Color	Cool 4 deg C	M&MM01	O41	01/09/2024	SM2120 B

Date/Time 01/09/24 12:35
Raw Sample Received by: 12/201
Raw Sample Relinquished by: 786001

Date/Time 01/09/24 13:20
Raw Sample Received by: 786007
Raw Sample Relinquished by: 12/201