



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129080

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/09/2024	08:45
2	P1055-02		1	12	56.00	132.8	132.5	01/09/2024	09:15
3	P1055-02DUP		1	11	56.00	132.8	132.5	01/09/2024	09:45
4	P1055-04		1	13	53.00	127.4	127.1	01/09/2024	10:15
5	P1055-06		1	3	7.00	44.6	44.3	01/09/2024	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

65129080

WorkList Name : fp-0109

WorkList ID : 177036

Department : Wet-Chemistry

Date : 01-09-2024 08:29:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1055-02	TOTE-1221	Solid	Flash Point	Cool 4 deg C	PSEG03	O31	01/05/2024	1010B
P1055-04	TOTE-1222	Solid	Flash Point	Cool 4 deg C	PSEG03	O31	01/05/2024	1010B
P1055-06	TOTE-1223	Solid	Flash Point	Cool 4 deg C	PSEG03	O31	01/05/2024	1010B

Date/Time 01/09/2024 08:35
Raw Sample Received by: RM (w.c.)
Raw Sample Relinquished by: RM (w.c.)

Date/Time 01/09/2024 11:20
Raw Sample Received by: RM (w.c.)
Raw Sample Relinquished by: RM (w.c.)