


LA129082

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.4525e-004 *slope*

C= -6.6206e-002 *y-intercept*

Rho= 0.9982260

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.00	0.000	-0.058	-0.058	32	0.000	32					2
0.05	0.050					442	0				-3
0.20	0.200	0.195	-0.005	1066	0.0 %	1066					1
2.50	2.500	2.537	0.037	10615	0.0 %	10615					-2
5.00	5.000	4.917	-0.083	20318	0.0 %	20318					6
7.50	7.500	7.923	0.423	32578	0.0 %	32578					-3
10.0	10.000	9.686	-0.314	39765	0.0 %	39765					

LB129082

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	32	-		0 SFAM01.1	PPB	1/9/2024 15:47	Std
	0.2 S01	1066	-		0.2 SFAM01.1	PPB	1/9/2024 15:49	Std
	2.5 S02	10615	-		2.5 SFAM01.1	PPB	1/9/2024 15:51	Std
	5 S03	20318	-		5 SFAM01.1	PPB	1/9/2024 15:54	Std
	7.5 S04	32578	-		7.5 SFAM01.1	PPB	1/9/2024 16:04	Std
	10 S05	39765	-		10 SFAM01.1	PPB	1/9/2024 16:12	Std
ICV058	ICV058	17299	4.1763 -		SFAM01.1	PPB	1/9/2024 16:15	SMPL
ICB058	ICB058	15	-0.0625 -		SFAM01.1	PPB	1/9/2024 16:20	SMPL
CCV046	CCV046	21135	5.1171 -		SFAM01.1	PPB	1/9/2024 16:22	SMPL
CCB046	CCB046	10	-0.0638 -		SFAM01.1	PPB	1/9/2024 16:24	SMPL
PB158332BL	PBW332	253	-0.0042 -		SFAM01.1	PPB	1/9/2024 16:26	SMPL
O5923-01	EOWS8	4579	1.0568 -		SFAM01.1	PPB	1/9/2024 16:29	SMPL
O5923-02	EOWT6	1386	0.2737 -		SFAM01.1	PPB	1/9/2024 16:31	SMPL
O5923-03	EOWW4	2924	0.6509 -		SFAM01.1	PPB	1/9/2024 16:33	SMPL
O5923-04	EOWW4D	2631	0.579 -		SFAM01.1	PPB	1/9/2024 16:35	SMPL
O5923-05	EOWW4S	7446	1.7599 -		SFAM01.1	PPB	1/9/2024 16:38	SMPL
O5923-06	EOWX2	523	0.0621 -		SFAM01.1	PPB	1/9/2024 16:40	SMPL
O5923-07	E0X02	404	0.0329 -		SFAM01.1	PPB	1/9/2024 16:42	SMPL
O5923-08	EOWW5	534	0.0648 -		SFAM01.1	PPB	1/9/2024 16:45	SMPL
O5923-09	E0X03	580	0.076 -		SFAM01.1	PPB	1/9/2024 16:47	SMPL
O5923-10	E0X23	379	0.0267 -		SFAM01.1	PPB	1/9/2024 16:49	SMPL
O5923-11	E0XY2	313	0.0106 -		SFAM01.1	PPB	1/9/2024 16:51	SMPL
O5923-12	E0XY3	236	-0.0083 -		SFAM01.1	PPB	1/9/2024 16:54	SMPL
O5923-13	EOWS9	242	-0.0069 -		SFAM01.1	PPB	1/9/2024 16:56	SMPL
O5923-14	E0X04	231	-0.0096 -		SFAM01.1	PPB	1/9/2024 16:58	SMPL
O5923-15	E0X24	232	-0.0093 -		SFAM01.1	PPB	1/9/2024 17:00	SMPL
PB158333BL	PBW333	615	0.0846 -		SFAM01.1	PPB	1/9/2024 17:03	SMPL
O5924-01	EOWX3	569	0.0733 -		SFAM01.1	PPB	1/9/2024 17:05	SMPL
O5924-02	EOWX8	515	0.0601 -		SFAM01.1	PPB	1/9/2024 17:07	SMPL
O5924-03	EOWZ4	1969	0.4167 -		SFAM01.1	PPB	1/9/2024 17:09	SMPL
CCV047	CCV047	22803	5.5261 -		SFAM01.1	PPB	1/9/2024 17:12	SMPL
CCB047	CCB047	158	-0.0275 -		SFAM01.1	PPB	1/9/2024 17:14	SMPL
O5924-04	EOWZ4D	1885	0.3961 -		SFAM01.1	PPB	1/9/2024 17:16	SMPL
O5924-05	EOWZ4S	7696	1.8212 -		SFAM01.1	PPB	1/9/2024 17:19	SMPL
O5924-06	EOWY6	31	-0.0586 -		SFAM01.1	PPB	1/9/2024 17:21	SMPL
O5924-07	E0X12	38	-0.0569 -		SFAM01.1	PPB	1/9/2024 17:23	SMPL
O5924-08	E0X32	42	-0.0559 -		SFAM01.1	PPB	1/9/2024 17:25	SMPL
O5924-09	EOWZ5	108	-0.0397 -		SFAM01.1	PPB	1/9/2024 17:28	SMPL
O5924-10	E0X13	73	-0.0483 -		SFAM01.1	PPB	1/9/2024 17:30	SMPL
O5924-11	E0X33	102	-0.0412 -		SFAM01.1	PPB	1/9/2024 17:32	SMPL
O5924-12	EOWX9	101	-0.0414 -		SFAM01.1	PPB	1/9/2024 17:34	SMPL
O5924-13	E0X14	136	-0.0329 -		SFAM01.1	PPB	1/9/2024 17:37	SMPL
O5924-14	E0X34	91	-0.0439 -		SFAM01.1	PPB	1/9/2024 17:39	SMPL
CCV048	CCV048	22664	5.492 -		SFAM01.1	PPB	1/9/2024 17:41	SMPL
CCB048	CCB048	270	0 -		SFAM01.1	PPB	1/9/2024 17:43	SMPL