

LB129085


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

63924



A= 0.0000e+000

B= 1.5981e-004

C= -1.1795e-001

Rho= 0.9994213

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.069	0.069	1173	0.000	1173					-
0.2	0.200	0.204	0.004	2015	0.0 %	2015					2
2.5	2.500	2.381	-0.119	15635	0.0 %	15635					-5
5.0	5.000	5.154	0.154	32988	0.0 %	32988					3
7.5	7.500	7.294	-0.206	46383	0.0 %	46383					-3
10.0	10.000	10.098	0.098	63924	0.0 %	63924					1

LB129085 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Jan 2024 08:10:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	4.0	1173	-	0.000	7471B	PPB	10 Jan 2024 08:13:01	S	Std
0.2 - 1	50.2	2015	-	0.200	7471B	PPB	10 Jan 2024 08:15:18	S	Std
2.5 - 1	52.5	15635	-	2.500	7471B	PPB	10 Jan 2024 08:17:35	S	Std
5.0 - 1	55	32988	-	5.000	7471B	PPB	10 Jan 2024 08:19:53	S	Std
7.5 - 1	57.5	46383	-	7.500	7471B	PPB	10 Jan 2024 08:22:13	S	Std
10.0 - 1	59.6	63924	-	10.000	7471B	PPB	10 Jan 2024 08:24:40	S	Std
ICV59 - 1	ICV59	26649	4.1408	-	7471B	PPB	10 Jan 2024 08:27:41	U	SMPL
ICB59 - 1	ICB59	812	0.0118	-	7471B	PPB	10 Jan 2024 08:29:58	U	SMPL
CCV49 - 1	CCV49	33528	5.2401	-	7471B	PPB	10 Jan 2024 08:32:21	U	SMPL
CCB49 - 1	CCB49	780	0.0067	-	7471B	PPB	10 Jan 2024 08:34:36	U	SMPL
CRA - 1	CRA	1950	0.1937	-	7471B	PPB	10 Jan 2024 08:37:00	U	SMPL
HighStd - 1	HighStd	63690	10.0602	-	7471B	PPB	10 Jan 2024 08:39:15	U	SMPL
ChkStd - 1	ChkStd	44759	7.0349	-	7471B	PPB	10 Jan 2024 08:41:30	U	SMPL
PB158328BL - 1	PBS	726	-0.0019	-	7471B	PPB	10 Jan 2024 08:46:47	U	SMPL
PB158328BS - 1	LCSS	26019	4.0401	-	7471B	PPB	10 Jan 2024 08:49:03	U	SMPL
P1050-01 - 1	CHRT24029	7561	1.0904	-	7471B	PPB	10 Jan 2024 08:51:19	U	SMPL
P1051-01 - 1	340	1367	0.1005	-	7471B	PPB	10 Jan 2024 08:53:44	U	SMPL
P1054-01 - 1	TR-06-01052024	2890	0.3439	-	7471B	PPB	10 Jan 2024 08:56:04	U	SMPL
P1055-07 - 1	SPILL-DEBRIS-COMP	615	-0.0197	-	7471B	PPB	10 Jan 2024 08:58:21	U	SMPL
P1056-01 - 1	MR-311P-0-2	21661	3.3436	-	7471B	PPB	10 Jan 2024 09:00:38	U	SMPL
CCV50 - 1	CCV50	33017	5.1584	-	7471B	PPB	10 Jan 2024 09:02:54	U	SMPL
CCB50 - 1	CCB50	868	0.0208	-	7471B	PPB	10 Jan 2024 09:05:16	U	SMPL
P1056-03 - 1	MR-311P-30-32	806	0.0109	-	7471B	PPB	10 Jan 2024 09:07:39	U	SMPL
P1067-01 - 1	SP-A-COMP(0-8)	1540	0.1281	-	7471B	PPB	10 Jan 2024 09:09:55	U	SMPL
P1067-01DUP - 1	SP-A-COMP(0-8)DUP	1594	0.1368	-	7471B	PPB	10 Jan 2024 09:12:10	U	SMPL
P1067-01MS - 1	SP-A-COMP(0-8)MS	30570	4.7674	-	7471B	PPB	10 Jan 2024 09:14:26	U	SMPL
P1067-01MSD - 1	SP-A-COMP(0-8)MSD	29603	4.6128	-	7471B	PPB	10 Jan 2024 09:16:42	U	SMPL
P1067-04 - 1	SP-B-COMP(0-8)	4937	0.6710	-	7471B	PPB	10 Jan 2024 09:19:07	U	SMPL
P1067-07 - 1	SP-C-COMP(0-8)	296497	47.2646	-	7471B	PPB	10 Jan 2024 09:21:32	U	SMPL
P1068-01 - 1	CL-01-010824	3406	0.4264	-	7471B	PPB	10 Jan 2024 09:23:51	U	SMPL
P1069-01 - 1	EO-01-010824	2255	0.2424	-	7471B	PPB	10 Jan 2024 09:26:37	U	SMPL
P1070-01 - 1	HD-01-010824	3819	0.4924	-	7471B	PPB	10 Jan 2024 09:28:54	U	SMPL
CCV51 - 1	CCV51	33412	5.2215	-	7471B	PPB	10 Jan 2024 09:31:11	U	SMPL
CCB51 - 1	CCB51	923	0.0295	-	7471B	PPB	10 Jan 2024 09:33:28	U	SMPL
P1073-01 - 1	69-2-TP	1312	0.0917	-	7471B	PPB	10 Jan 2024 09:35:54	U	SMPL
P1067-01LX5 - 1		263	-0.0759	-	7471B	PPB	10 Jan 2024 09:38:09	U	SMPL
P1067-01A - 1		28936	4.5062	-	7471B	PPB	10 Jan 2024 09:40:25	U	SMPL
P1067-07DLX10 - 1		31048	4.8438	-	7471B	PPB	10 Jan 2024 09:42:40	U	SMPL
CCV52 - 1	CCV52	33558	5.2449	-	7471B	PPB	10 Jan 2024 09:45:03	U	SMPL
CCB52 - 1	CCB52	927	0.0302	-	7471B	PPB	10 Jan 2024 09:47:26	U	SMPL