



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129086

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/10/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.01	01/10/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.03	01/10/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.01	01/10/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.02	01/10/2024	08:20
6	P1083-01	1	Water	NA	NA	20.1	6.60	01/10/2024	08:25
7	P1083-01DUP	1	Water	NA	NA	20.2	6.61	01/10/2024	08:26
8	P1083-03	1	Water	NA	NA	20.1	6.82	01/10/2024	08:35
9	P1084-01	1	Water	NA	NA	20.7	9.52	01/10/2024	08:40
10	CCV2	1	Water	NA	NA	20.2	12.02	01/10/2024	08:45

WORKLIST(Hardcopy Internal Chain)

129086

WorkList Name : ph p1084

WorkList ID : 177065

Department : Wet-Chemistry

Date : 01-10-2024 07:56:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1083-01	A GROUND-WATER-TOTE	Water	pH	Cool 4 deg C	PSEG03	O31	01/09/2024	9040C
P1083-03	A OILY-WATER-TOTE	Water	pH	Cool 4 deg C	PSEG03	O31	01/09/2024	9040C
P1084-01	A OWS-PROJECT-COMP	Water	pH	Cool 4 deg C	PSEG03	O31	01/09/2024	9040C

Date/Time 01/10/24 08:00

Raw Sample Received by: SP (QC)

Raw Sample Relinquished by: CP SM

Date/Time 01/10/24

Raw Sample Received by: CP SM

Raw Sample Relinquished by: SP (QC)