


LB129093

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.8924e-004 slope

C= 1.9588e-002 y-intercept

Rho= 0.9999583

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%1)
0.00	0.000	0.033	0.033	71	0.000	71					11
0.05	0.050					442	0				11
0.20	0.200	0.222	0.022	1067	0.0 %	1067					11
2.50	2.500	2.437	-0.063	12777	0.0 %	12777					11
5.00	5.000	4.997	-0.003	26301	0.0 %	26301					11
7.50	7.500	7.480	-0.020	39425	0.0 %	39425					11
10.0	10.000	10.031	0.031	52903	0.0 %	52903					11

LB129093

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	71	-		0 SFAM01.1	PPB	1/10/2024 11:34	Std
	0.2 S01	1067	-		0.2 SFAM01.1	PPB	1/10/2024 11:36	Std
	2.5 S02	12777	-		2.5 SFAM01.1	PPB	1/10/2024 11:39	Std
	5 S03	26301	-		5 SFAM01.1	PPB	1/10/2024 11:43	Std
	7.5 S04	39425	-		7.5 SFAM01.1	PPB	1/10/2024 11:45	Std
	10 S05	52903	-		10 SFAM01.1	PPB	1/10/2024 11:47	Std
ICV060	ICV060	20115	3.8261 -		SFAM01.1	PPB	1/10/2024 11:50	SMPL
ICB060	ICB060	-299	-0.037 -		SFAM01.1	PPB	1/10/2024 11:53	SMPL
CCV053	CCV053	26640	5.0609 -		SFAM01.1	PPB	1/10/2024 11:55	SMPL
CCB053	CCB053	-306	-0.0383 -		SFAM01.1	PPB	1/10/2024 11:57	SMPL
PB158351BL	PBW251	275	0.0716 -		SFAM01.1	PPB	1/10/2024 12:01	SMPL
PB158082TB	LEB251	192	0.0559 -		SFAM01.1	PPB	1/10/2024 12:06	SMPL
O5943-01	E0X84	1196	0.2459 -		SFAM01.1	PPB	1/10/2024 12:09	SMPL
O5943-02	E0XK1	997	0.2083 -		SFAM01.1	PPB	1/10/2024 12:11	SMPL
O5943-03	E0XY6	898	0.1895 -		SFAM01.1	PPB	1/10/2024 12:18	SMPL
O5943-04	E0XY6D	754	0.1623 -		SFAM01.1	PPB	1/10/2024 12:20	SMPL
O5943-05	E0XY6S	6391	1.229 -		SFAM01.1	PPB	1/10/2024 12:25	SMPL
CCV054	CCV054	26939	5.1175 -		SFAM01.1	PPB	1/10/2024 12:27	SMPL
CCB054	CCB054	-263	-0.0302 -		SFAM01.1	PPB	1/10/2024 12:30	SMPL