



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/11/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 01/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 01/11/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB129098

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1022-01	E0X91	1	1.13	8.60	9.73	8.82	89.4	
P1022-02	E0X95	2	1.15	8.39	9.54	8.77	90.8	
P1022-03	E0X96	3	1.18	8.66	9.84	9.00	90.3	
P1022-04	E0X97	4	1.19	8.53	9.72	8.79	89.1	
P1022-05	E0X98	5	1.13	8.44	9.57	7.77	78.7	
P1022-06	E0XN1	6	1.18	8.34	9.52	7.8	79.4	
P1022-07	E0X91	7	1.16	8.72	9.88	9.06	90.6	
P1022-08	E0X92	8	1.19	8.52	9.71	8.88	90.3	
P1022-09	E0X93	9	1.14	8.83	9.97	7.76	75.0	
P1022-10	E0X93D	10	1.14	8.83	9.97	7.76	75.0	
P1022-11	E0X93S	11	1.14	8.83	9.97	7.76	75.0	
P1022-12	E0X79	12	1.14	8.83	9.97	9.28	92.2	
P1022-13	E0X80	13	1.17	8.49	9.66	8.93	91.4	
P1022-14	E0X81	14	1.18	8.57	9.75	8.8	88.9	
P1022-15	E0X82	15	1.18	8.61	9.79	8.96	90.4	
P1022-16	E0XN2	16	1.12	8.44	9.56	8.94	92.7	
P1022-20	E0XH6	17	1.17	8.71	9.88	8.97	89.6	
P1022-21	E0XH7	18	1.19	8.42	9.61	8.03	81.2	
P1022-22	E0XX8	19	1.17	8.57	9.74	7.8	77.4	
P1022-23	E0X75	20	1.13	8.55	9.68	9.11	93.3	
P1022-24	E0X76	21	1.16	8.74	9.9	9.35	93.7	
P1022-25	E0X77	22	1.11	8.48	9.59	9.12	94.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$