

LB129100


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53490



A= 0.0000e+000

B= 1.8704e-004 slope

C= -1.0640e-001 y-intercept

Rho= 0.9996668

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.00	0.000	-0.055	-0.055	273	0.000	273					11
0.05	0.050					442	0				1
0.20	0.200	0.192	-0.008	1596	0.0 %	1596					2
2.50	2.500	2.598	0.098	14460	0.0 %	14460					4
5.00	5.000	4.913	-0.087	26835	0.0 %	26835					2
7.50	7.500	7.654	0.154	41491	0.0 %	41491					2
10.0	10.000	9.898	-0.102	53490	0.0 %	53490					1

LB129100

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	273	-		0 SFAM01.1	PPB	1/10/2024 14:00	Std
	0.2 S01	1596	-		0.2 SFAM01.1	PPB	1/10/2024 14:02	Std
	2.5 S02	14460	-		2.5 SFAM01.1	PPB	1/10/2024 14:04	Std
	5 S03	26835	-		5 SFAM01.1	PPB	1/10/2024 14:07	Std
	7.5 S04	41491	-		7.5 SFAM01.1	PPB	1/10/2024 14:09	Std
	10 S05	53490	-		10 SFAM01.1	PPB	1/10/2024 14:14	Std
ICV061	ICV061	21049	3.8306 -		SFAM01.1	PPB	1/10/2024 14:17	SMPL
ICB061	ICB061	316	-0.0473 -		SFAM01.1	PPB	1/10/2024 14:19	SMPL
CCV055	CCV055	27426	5.0233 -		SFAM01.1	PPB	1/10/2024 14:21	SMPL
CCB055	CCB055	312	-0.048 -		SFAM01.1	PPB	1/10/2024 14:24	SMPL
PB158352BL	PBS352	604	0.0066 -		SFAM01.1	PPB	1/10/2024 14:26	SMPL
O5944-01	E0X90	3242	0.5 -		SFAM01.1	PPB	1/10/2024 14:28	SMPL
O5944-02	E0XN0	831	0.049 -		SFAM01.1	PPB	1/10/2024 14:30	SMPL
O5944-03	E0XX9	3390	0.5277 -		SFAM01.1	PPB	1/10/2024 14:33	SMPL
O5944-04	E0XY0	1043	0.0887 -		SFAM01.1	PPB	1/10/2024 14:35	SMPL
O5944-05	E0X55	1586	0.1902 -		SFAM01.1	PPB	1/10/2024 14:37	SMPL
O5944-06	E0X56	914	0.0646 -		SFAM01.1	PPB	1/10/2024 14:39	SMPL
O5944-08	E0XY1	983	0.0775 -		SFAM01.1	PPB	1/10/2024 14:42	SMPL
O5944-09	E0X57	1155	0.1096 -		SFAM01.1	PPB	1/10/2024 14:44	SMPL
O5944-10	E0X57D	1162	0.1109 -		SFAM01.1	PPB	1/10/2024 14:46	SMPL
O5944-11	E0X57S	17272	3.1241 -		SFAM01.1	PPB	1/10/2024 14:49	SMPL
O5944-12	E0X58	10679	1.891 -		SFAM01.1	PPB	1/10/2024 14:51	SMPL
O5944-13	E0XY6	1078	0.0952 -		SFAM01.1	PPB	1/10/2024 14:53	SMPL
O5944-14	E0XY7	1335	0.1433 -		SFAM01.1	PPB	1/10/2024 14:55	SMPL
CCV056	CCV056	28153	5.1593 -		SFAM01.1	PPB	1/10/2024 14:58	SMPL
CCB056	CCB056	428	-0.0264 -		SFAM01.1	PPB	1/10/2024 15:00	SMPL