



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129103

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/10/2024	15:30
2	CAL2	1	Water	NA	NA	20.3	7.01	01/10/2024	15:31
3	CAL3	1	Water	NA	NA	20.3	10.02	01/10/2024	15:32
4	ICV	1	Water	NA	NA	20.3	6.98	01/10/2024	15:38
5	CCV1	1	Water	NA	NA	20.3	2.01	01/10/2024	15:40
6	P1098-01	1	Water	NA	NA	20.5	6.02	01/10/2024	15:50
7	P1098-01DUP	1	Water	NA	NA	20.6	6.03	01/10/2024	15:51
8	P1098-04	1	Water	NA	NA	20.3	6.09	01/10/2024	16:00
9	CCV2	1	Water	NA	NA	20.3	12.02	01/10/2024	16:10

129103

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1098 WorkList ID : 177093 Department : Wet-Chemistry Date : 01-10-2024 15:14:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1098-01	AUD-1310	Water	pH	Cool 4 deg C	PSEG03	O31	01/10/2024	9040C
P1098-04	AUD-1311	Water	pH	Cool 4 deg C	PSEG03	O31	01/10/2024	9040C

Date/Time 01-10-24 15:20
Raw Sample Received by: 801901
Raw Sample Relinquished by: [Signature]

Date/Time 01-10-24 17:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 801901