



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129115

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	01/11/2024	08:30
2	P1098-01		1	12	100.00	>212.0	>212.0	01/11/2024	09:00
3	P1098-01DUP		1	14	100.00	>212.0	>212.0	01/11/2024	09:30
4	P1098-04		1	13	100.00	>212.0	>212.0	01/11/2024	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB129115

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint01/11/2024

WorkList ID : 177106

Department : Wet-Chemistry

Date : 01-11-2024 08:10:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1098-01	AUD-1310	Water	Flash Point	Cool 4 deg C	PSEG03	O31	01/10/2024	1010B
P1098-04	AUD-1311	Water	Flash Point	Cool 4 deg C	PSEG03	O31	01/10/2024	1010B

Date/Time 01/11/2024 8:15
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.

Date/Time 01/11/2024 10:30
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.