

SFAM01.1

Linear

$$\mu \text{ Abs.} :$$

51051

A= 0.0000e+000

B= 1.9643e-004 slope

C = 3.2135e-002 *y-intercept*

Rho= 0.9998989

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	OTCD
0.00	0.000	0.061	0.061	148	0.000	148					11
0.05	0.050					442	0				13
0.20	0.200	0.226	0.026	987	0.0 %	987					-3
2.50	2.500	2.434	-0.066	12229	0.0 %	12229					-1
5.00	5.000	4.938	-0.062	24978	0.0 %	24978					0
7.50	7.500	7.480	-0.020	37918	0.0 %	37918					
10.0	10.000	10.060	0.060	51051	0.0 %	51051					

$$\begin{array}{r} 1000 \\ -1113 \\ -3101 \\ \hline \end{array}$$

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	148	-		0 SFAM01.1	PPB	1/11/2024 14:37	Std
	0.2 S01	987	-		0.2 SFAM01.1	PPB	1/11/2024 14:39	Std
	2.5 S02	12229	-		2.5 SFAM01.1	PPB	1/11/2024 14:41	Std
	5 S03	24978	-		5 SFAM01.1	PPB	1/11/2024 14:43	Std
	7.5 S04	37918	-		7.5 SFAM01.1	PPB	1/11/2024 14:46	Std
	10 S05	51051	-		10 SFAM01.1	PPB	1/11/2024 14:48	Std
ICV063	ICV063	19536	3.8695 -		SFAM01.1	PPB	1/11/2024 14:51	SMPL
ICB063	ICB063	-102	0.0121 -		SFAM01.1	PPB	1/11/2024 14:53	SMPL
CCV061	CCV061	25097	4.9618 -		SFAM01.1	PPB	1/11/2024 14:55	SMPL
CCB061	CCB061	-93	0.0139 -		SFAM01.1	PPB	1/11/2024 14:58	SMPL
PB158385BL	PBW385	86	0.049 -		SFAM01.1	PPB	1/11/2024 15:00	SMPL
O6036-01	YBRR0	473	0.125 -		SFAM01.1	PPB	1/11/2024 15:02	SMPL
O6036-02	YBRR0D	642	0.1582 -		SFAM01.1	PPB	1/11/2024 15:05	SMPL
O6036-03	YBRR0S	5589	1.13 -		SFAM01.1	PPB	1/11/2024 15:07	SMPL
O6036-04	YBRR1	1122	0.2525 -		SFAM01.1	PPB	1/11/2024 15:09	SMPL
O6036-05	YBRR2	1891	0.4036 -		SFAM01.1	PPB	1/11/2024 15:11	SMPL
O6036-06	YBRR3	3399	0.6998 -		SFAM01.1	PPB	1/11/2024 15:14	SMPL
O6036-07	YBRR4	1905	0.4063 -		SFAM01.1	PPB	1/11/2024 15:16	SMPL
O6036-08	YBRR5	1139	0.2559 -		SFAM01.1	PPB	1/11/2024 15:18	SMPL
PB158386BL	PBW386	71	0.0461 -		SFAM01.1	PPB	1/11/2024 15:21	SMPL
P1023-01	EOXY9	119	0.0555 -		SFAM01.1	PPB	1/11/2024 15:23	SMPL
P1023-02	EOXZ0	143	0.0602 -		SFAM01.1	PPB	1/11/2024 15:25	SMPL
P1023-03	EOX05	166	0.0647 -		SFAM01.1	PPB	1/11/2024 15:27	SMPL
P1023-04	EOX25	247	0.0807 -		SFAM01.1	PPB	1/11/2024 15:30	SMPL
P1023-05	EOXZ2	256	0.0824 -		SFAM01.1	PPB	1/11/2024 15:32	SMPL
P1023-06	EOX06	114	0.0545 -		SFAM01.1	PPB	1/11/2024 15:34	SMPL
P1023-07	EOXZ3	236	0.0785 -		SFAM01.1	PPB	1/11/2024 15:36	SMPL
P1023-08	EOXZ6	261	0.0834 -		SFAM01.1	PPB	1/11/2024 15:39	SMPL
P1023-09	EOX26	217	0.0748 -		SFAM01.1	PPB	1/11/2024 15:41	SMPL
P1023-10	EOXZ8	287	0.0885 -		SFAM01.1	PPB	1/11/2024 15:43	SMPL
CCV062	CCV062	24227	4.7909 -		SFAM01.1	PPB	1/11/2024 15:45	SMPL
CCB062	CCB062	-18	0.0286 -		SFAM01.1	PPB	1/11/2024 15:48	SMPL
P1023-11	EOXZ9	88	0.0494 -		SFAM01.1	PPB	1/11/2024 15:50	SMPL
P1023-12	EOY01	72	0.0463 -		SFAM01.1	PPB	1/11/2024 15:52	SMPL
P1043-01	EOX15	64	0.0447 -		SFAM01.1	PPB	1/11/2024 15:55	SMPL
P1043-02	EOX35	85	0.0488 -		SFAM01.1	PPB	1/11/2024 15:57	SMPL
P1043-03	EOX16	79	0.0477 -		SFAM01.1	PPB	1/11/2024 15:59	SMPL
P1043-04	EOX36	116	0.0549 -		SFAM01.1	PPB	1/11/2024 16:01	SMPL
CCV063	CCV063	24510	4.8465 -		SFAM01.1	PPB	1/11/2024 16:04	SMPL
CCB063	CCB063	-28	0.0266 -		SFAM01.1	PPB	1/11/2024 16:06	SMPL