



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB129122

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/11/2024	15:15
2	CAL2	1	Water	NA	NA	20.3	7.00	01/11/2024	15:16
3	CAL3	1	Water	NA	NA	20.3	10.02	01/11/2024	15:19
4	ICV	1	Water	NA	NA	20.5	7.01	01/11/2024	15:22
5	CCV1	1	Water	NA	NA	20.6	2.02	01/11/2024	15:25
6	P1105-01	1	Water	NA	NA	20.4	4.27	01/11/2024	15:30
7	P1105-01DUP	1	Water	NA	NA	20.5	4.28	01/11/2024	15:31
8	P1105-02	1	Water	NA	NA	20.4	5.69	01/11/2024	15:35
9	P1105-03	1	Water	NA	NA	20.7	5.61	01/11/2024	15:39
10	P1105-04	1	Water	NA	NA	20.7	5.76	01/11/2024	15:44
11	P1105-05	1	Water	NA	NA	20.6	5.62	01/11/2024	15:50
12	P1106-01	1	Water	NA	NA	20.5	5.39	01/11/2024	15:52
13	P1109-01	1	Water	NA	NA	20.5	4.39	01/11/2024	16:00
14	P1110-01	1	Water	NA	NA	20.7	4.31	01/11/2024	16:10
15	P1110-02	1	Water	NA	NA	20.6	4.26	01/11/2024	16:15
16	CCV2	1	Water	NA	NA	20.3	12.02	01/11/2024	16:20
17	P1110-03	1	Water	NA	NA	20.7	4.00	01/11/2024	16:25
18	CCV3	1	Water	NA	NA	20.3	2.01	01/11/2024	16:26

WORKLIST(Hardcopy Internal Chain)

129122

WorkList Name : PH P1106

WorkList ID : 177118

Department : Wet-Chemistry

Date : 01-11-2024 15:02:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1105-01	60126	Water	pH	Cool 4 deg C	PSEG03	O31	01/11/2024	9040C
P1105-02	60127	Water	pH	Cool 4 deg C	PSEG03	O31	01/11/2024	9040C
P1105-03	60128	Water	pH	Cool 4 deg C	PSEG03	O31	01/11/2024	9040C
P1105-04	60129	Water	pH	Cool 4 deg C	PSEG03	O31	01/11/2024	9040C
P1105-05	60130	Water	pH	Cool 4 deg C	PSEG03	O31	01/11/2024	9040C
P1106-01	34495	Water	pH	Cool 4 deg C	PSEG03	O31	01/11/2024	9040C
P1109-01	CELL-A	Water	pH	Cool 4 deg C	PSEG03	O11	01/11/2024	9040C
P1110-01	CELL-A	Water	pH	Cool 4 deg C	STER10	O11	01/08/2024	9040C
P1110-02	CELL-B/C	Water	pH	Cool 4 deg C	STER07	O11	12/13/2023	9040C
P1110-03	CELL-D	Water	pH	Cool 4 deg C	STER07	O11	12/13/2023	9040C

Date/Time

01-11-24 15:10

Raw Sample Received by:

SP

Raw Sample Relinquished by:

SP

Date/Time

01-11-24

Raw Sample Received by:

SP

Raw Sample Relinquished by:

SP