


LB129131

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.6600e-004

C= 4.2346e-002

Rho= 0.9999206

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	a/o)
0.0	0.000	0.048	0.048	34	0.000	34					1
0.2	0.200	0.225	0.025	1100	0.0 %	1100					13
2.5	2.500	2.408	-0.092	14253	0.0 %	14253					24
5.0	5.000	4.978	-0.022	29736	0.0 %	29736					0
7.5	7.500	7.534	0.034	45129	0.0 %	45129					0
10.0	10.000	10.007	0.007	60027	0.0 %	60027					0

LB129131 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Jan 2024 11:14:52

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	12 Jan 2024 11:32:06	S	Std
0.2 - 1	50.2	1100	-	0.2000	7471B	PPB	12 Jan 2024 11:34:24	S	Std
2.5 - 1	52.5	14253	-	2.5000	7471B	PPB	12 Jan 2024 11:36:41	S	Std
5.0 - 1	55	29736	-	5.0000	7471B	PPB	12 Jan 2024 11:38:59	S	Std
7.5 - 1	57.5	45129	-	7.5000	7471B	PPB	12 Jan 2024 11:41:18	S	Std
10.0 - 1	50	60027	-	10.0000	7471B	PPB	12 Jan 2024 11:46:30	S	Std
ICV64 - 1	ICV64	23485	3.9408	-	7471B	PPB	12 Jan 2024 11:51:53	U	SMPL
ICB64 - 1	ICB64	-250	0.0008	-	7471B	PPB	12 Jan 2024 11:54:10	U	SMPL
CCV64 - 1	CCV64	31482	5.2683	-	7471B	PPB	12 Jan 2024 11:56:36	U	SMPL
CCB64 - 1	CCB64	-259	-0.0006	-	7471B	PPB	12 Jan 2024 11:58:51	U	SMPL
CRA - 1	CRA	1042	0.2153	-	7471B	PPB	12 Jan 2024 12:01:17	U	SMPL
HighStd - 1	HighStd	60012	10.0042	-	7471B	PPB	12 Jan 2024 12:03:32	U	SMPL
ChkStd - 1	ChkStd	42287	7.0619	-	7471B	PPB	12 Jan 2024 12:05:47	U	SMPL
PB158406BL - 1	PBS	-319	-0.0106	-	7471B	PPB	12 Jan 2024 12:08:21	U	SMPL
PB158406BS - 1	LCSS	25073	4.2044	-	7471B	PPB	12 Jan 2024 12:10:50	U	SMPL
P1085-01 - 1	TP-1	4966	0.8667	-	7471B	PPB	12 Jan 2024 12:13:06	U	SMPL
P1085-03 - 1	TP-2	15236	2.5715	-	7471B	PPB	12 Jan 2024 12:15:31	U	SMPL
P1085-05 - 1	TP-3	25484	4.2726	-	7471B	PPB	12 Jan 2024 12:17:50	U	SMPL
P1085-07 - 1	TP-4	2360	0.4341	-	7471B	PPB	12 Jan 2024 12:20:10	U	SMPL
P1086-01 - 1	215	878	0.1881	-	7471B	PPB	12 Jan 2024 12:22:35	U	SMPL
CCV65 - 1	CCV65	31338	5.2444	-	7471B	PPB	12 Jan 2024 12:24:51	U	SMPL
CCB65 - 1	CCB65	-203	0.0086	-	7471B	PPB	12 Jan 2024 12:27:07	U	SMPL
P1086-02 - 1	332	1495	0.2905	-	7471B	PPB	12 Jan 2024 12:29:34	U	SMPL
P1097-01 - 1	MH-G	1697	0.3240	-	7471B	PPB	12 Jan 2024 12:36:08	U	SMPL
P1097-01DUP - 1	MH-GDUP	1596	0.3073	-	7471B	PPB	12 Jan 2024 12:38:23	U	SMPL
P1097-01MS - 1	MH-GMS	27802	4.6574	-	7471B	PPB	12 Jan 2024 12:40:40	U	SMPL
P1097-01MSD - 1	MH-GMSD	27579	4.6204	-	7471B	PPB	12 Jan 2024 12:42:56	U	SMPL
P1100-01 - 1	COMP-1	5506	0.9563	-	7471B	PPB	12 Jan 2024 12:45:19	U	SMPL
P1100-02 - 1	COMP-2	16640	2.8045	-	7471B	PPB	12 Jan 2024 12:47:44	U	SMPL
P1100-03 - 1	COMP-3	4589	0.8041	-	7471B	PPB	12 Jan 2024 12:50:03	U	SMPL
P1104-01 - 1	TP-1	2493	0.4562	-	7471B	PPB	12 Jan 2024 12:52:24	U	SMPL
P1104-05 - 1	TP-2	388	0.1068	-	7471B	PPB	12 Jan 2024 12:54:42	U	SMPL
CCV66 - 1	CCV66	31636	5.2938	-	7471B	PPB	12 Jan 2024 12:56:58	U	SMPL
CCB66 - 1	CCB66	-193	0.0103	-	7471B	PPB	12 Jan 2024 12:59:14	U	SMPL
PB158407BL - 1	PBS	13	0.0445	-	7471B	PPB	12 Jan 2024 13:01:38	U	SMPL
PB158407BS - 1	LCSS	25758	4.3181	-	7471B	PPB	12 Jan 2024 13:03:54	U	SMPL
P1107-01 - 1	OK-01-11124	2103	0.3914	-	7471B	PPB	12 Jan 2024 13:06:09	U	SMPL
P1117-01 - 1	S1-COMPOSITE	988	0.2064	-	7471B	PPB	12 Jan 2024 13:08:34	U	SMPL
P1117-07 - 1	S2-COMPOSITE	1103	0.2254	-	7471B	PPB	12 Jan 2024 13:10:51	U	SMPL
P1118-01 - 1	WC-1	5350	0.9304	-	7471B	PPB	12 Jan 2024 13:13:07	U	SMPL
P1118-05 - 1	WC-2	6336	1.0941	-	7471B	PPB	12 Jan 2024 13:15:24	U	SMPL
P1119-01 - 1	MH-J	278	0.0885	-	7471B	PPB	12 Jan 2024 13:17:43	U	SMPL
P1119-01DUP - 1	MH-JDUP	362	0.1024	-	7471B	PPB	12 Jan 2024 13:23:09	U	SMPL
P1119-01MS - 1	MH-JMS	26628	4.4625	-	7471B	PPB	12 Jan 2024 13:25:25	U	SMPL
CCV67 - 1	CCV67	31480	5.2679	-	7471B	PPB	12 Jan 2024 13:27:41	U	SMPL
CCB67 - 1	CCB67	-200	0.0091	-	7471B	PPB	12 Jan 2024 13:30:04	U	SMPL
P1119-01MSD - 1	MH-JMSD	25476	4.2713	-	7471B	PPB	12 Jan 2024 13:32:28	U	SMPL
P1123-01 - 1	BASIN-5-TRUCK-STOP-TOP	454	0.1177	-	7471B	PPB	12 Jan 2024 13:34:43	U	SMPL
P1123-02 - 1	MEDIAN-WEST-SIDE-HOWARD-BLVD	3515	0.6258	-	7471B	PPB	12 Jan 2024 13:37:06	U	SMPL
P1123-03 - 1	MEDIAN-EAST-SIDE-HOWARD-BLVD	5044	0.8796	-	7471B	PPB	12 Jan 2024 13:39:22	U	SMPL
P1123-04 - 1	BASIN-4-RAMP-T	1816	0.3438	-	7471B	PPB	12 Jan 2024 13:41:40	U	SMPL
P1123-05 - 1	BASIN-5-TRUCK-STOP-BOTTOM	2514	0.4597	-	7471B	PPB	12 Jan 2024 13:43:58	U	SMPL
P1097-01LX5 - 1		322	0.0958	-	7471B	PPB	12 Jan 2024 13:46:16	U	SMPL
P1097-01A - 1		27876	4.6697	-	7471B	PPB	12 Jan 2024 13:48:40	U	SMPL
P1119-01LX5 - 1		-87	0.0279	-	7471B	PPB	12 Jan 2024 13:50:58	U	SMPL
P1119-01A - 1		25920	4.3450	-	7471B	PPB	12 Jan 2024 13:53:23	U	SMPL
CCV68 - 1	CCV68	30757	5.1479	-	7471B	PPB	12 Jan 2024 13:55:40	U	SMPL
CCB68 - 1	CCB68	-193	0.0103	-	7471B	PPB	12 Jan 2024 13:58:03	U	SMPL