



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129135

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/12/2024	13:10
2	CAL2	1	Water	NA	NA	20.3	7.00	01/12/2024	13:11
3	CAL3	1	Water	NA	NA	20.4	10.02	01/12/2024	13:12
4	ICV	1	Water	NA	NA	20.3	7.01	01/12/2024	13:15
5	CCV1	1	Water	NA	NA	20.3	2.01	01/12/2024	13:20
6	P1065-01	1	Solid	20.02	20	24.3	11.95	01/12/2024	13:22
7	P1065-02	1	Solid	20.03	20	24.2	10.98	01/12/2024	13:25
8	P1065-03	1	Solid	20.01	20	24.7	9.73	01/12/2024	13:35
9	P1065-04	1	Solid	20.03	20	24.6	10.26	01/12/2024	13:40
10	P1065-05	1	Solid	20.03	20	24.6	10.26	01/12/2024	13:40
11	P1065-06	1	Solid	20.03	20	24.6	10.26	01/12/2024	13:40
12	P1065-07	1	Solid	20.05	20	24.4	10.30	01/12/2024	13:45
13	P1065-08	1	Solid	20.05	20	23.6	9.80	01/12/2024	13:50
14	P1065-10	1	Solid	20.02	20	23.6	10.18	01/12/2024	14:00
15	P1065-11	1	Solid	20.04	20	23.7	10.31	01/12/2024	14:05
16	CCV2	1	Water	NA	NA	20.3	12.02	01/12/2024	14:10
17	P1065-12	1	Solid	20.02	20	23.8	9.84	01/12/2024	14:15
18	P1065-13	1	Solid	20.01	20	23.8	11.63	01/12/2024	14:25
19	P1065-14	1	Solid	20.03	20	23.8	9.23	01/12/2024	14:35
20	P1065-15	1	Solid	20.04	20	23.3	9.87	01/12/2024	14:40

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	P1065-16	1	Solid	20.02	20	23.5	9.48	01/12/2024	14:50
22	P1065-17	1	Solid	20.03	20	23.3	9.49	01/12/2024	15:00
23	CCV3	1	Water	NA	NA	20.4	2.02	01/12/2024	15:05

WORKLIST(Hardcopy Internal Chain)

129135

WorkList Name : ph p1065

WorkList ID : 177151

Department : Wet-Chemistry

Date : 01-12-2024 12:34:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1065-01	E0X94	Solid	pH	Cool 4 deg C	USEP01	Q21	01/03/2024	9045D
P1065-02	E0XZ4	Solid	pH	Cool 4 deg C	USEP01	Q21	01/04/2024	9045D
P1065-03	E0XZ7	Solid	pH	Cool 4 deg C	USEP01	Q21	01/05/2024	9045D
P1065-04	E0XH5	Solid	pH	Cool 4 deg C	USEP01	Q21	01/05/2024	9045D
P1065-05	E0XH5D	Solid	pH	Cool 4 deg C	USEP01	Q21	01/05/2024	9045D
P1065-06	E0XH5S	Solid	pH	Cool 4 deg C	USEP01	Q21	01/05/2024	9045D
P1065-07	E0X78	Solid	pH	Cool 4 deg C	USEP01	Q21	01/05/2024	9045D
P1065-08	E0X67	Solid	pH	Cool 4 deg C	USEP01	Q21	01/09/2024	9045D
P1065-10	E0Y05	Solid	pH	Cool 4 deg C	USEP01	Q21	01/10/2024	9045D
P1065-11	E0X68	Solid	pH	Cool 4 deg C	USEP01	Q21	01/10/2024	9045D
P1065-12	E0X69	Solid	pH	Cool 4 deg C	USEP01	Q21	01/10/2024	9045D
P1065-13	E0X70	Solid	pH	Cool 4 deg C	USEP01	Q21	01/10/2024	9045D
P1065-14	E0Y06	Solid	pH	Cool 4 deg C	USEP01	Q21	01/10/2024	9045D
P1065-15	E0Y07	Solid	pH	Cool 4 deg C	USEP01	Q21	01/10/2024	9045D
P1065-16	E0X63	Solid	pH	Cool 4 deg C	USEP01	Q21	01/11/2024	9045D
P1065-17	E0Y11	Solid	pH	Cool 4 deg C	USEP01	Q21	01/11/2024	9045D

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: