



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129139

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	01/15/2024	08:20
2	P1105-01		1	14	100.00	>212.0	>212.0	01/15/2024	08:50
3	P1105-01DUP		1	12	100.00	>212.0	>212.0	01/15/2024	09:20
4	P1105-02		1	13	100.00	>212.0	>212.0	01/15/2024	09:50
5	P1105-03		1	14	100.00	>212.0	>212.0	01/15/2024	10:20
6	P1105-04		1	12	100.00	>212.0	>212.0	01/15/2024	10:50
7	P1105-05		1	14	100.00	>212.0	>212.0	01/15/2024	11:20
8	P1106-01		1	12	100.00	>212.0	>212.0	01/15/2024	11:50
9	P1133-01		1	13	100.00	>212.0	>212.0	01/15/2024	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129139

WORKLIST(Hardcopy Internal Chain)

WorkList Name : falshpoint01/12/24 WorkList ID : 177142 Department : Wet-Chemistry Date : 01-12-2024 09:44:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1105-01	60126	Water	Flash Point	Cool 4 deg C	PSEG03	O31	01/11/2024	1010B
P1105-02	60127	Water	Flash Point	Cool 4 deg C	PSEG03	O31	01/11/2024	1010B
P1105-03	60128	Water	Flash Point	Cool 4 deg C	PSEG03	O31	01/11/2024	1010B
P1105-04	60129	Water	Flash Point	Cool 4 deg C	PSEG03	O31	01/11/2024	1010B
P1105-05	60130	Water	Flash Point	Cool 4 deg C	PSEG03	O31	01/11/2024	1010B
P1106-01	34495	Water	Flash Point	Cool 4 deg C	PSEG03	O11	01/11/2024	1010B
P1133-01	TOTE-6579	Water	Flash Point	Cool 4 deg C	PSEG03	O41	01/12/2024	1010B

Date/Time 01/15/2024 8:40
Raw Sample Received by: N.P W-C
Raw Sample Relinquished by: N.P W-C

Date/Time 01/15/2024 13:00
Raw Sample Received by: N.P W-C
Raw Sample Relinquished by: N.P W-C