

LB129140


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46211



A= 0.0000e+000

B= 2.1491e-004 slope

C= -3.8212e-002 y-intercept

Rho= 0.9995586

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1)
0.00	0.000	-0.010	-0.010	133	0.000	133					=
0.05	0.050					442	0				-14
0.20	0.200	0.172	-0.028	976	0.0 %	976					3
2.50	2.500	2.581	0.081	12186	0.0 %	12186					-3
5.00	5.000	4.872	-0.128	22850	0.0 %	22850					3
7.50	7.500	7.692	0.192	35971	0.0 %	35971					-1
10.0	10.000	9.893	-0.107	46211	0.0 %	46211					

LB129140

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	133	-		0 SFAM01.1	PPB	1/12/2024 15:36	Std
	0.2 S01	976	-		0.2 SFAM01.1	PPB	1/12/2024 15:38	Std
	2.5 S02	12186	-		2.5 SFAM01.1	PPB	1/12/2024 15:41	Std
	5 S03	22850	-		5 SFAM01.1	PPB	1/12/2024 15:43	Std
	7.5 S04	35971	-		7.5 SFAM01.1	PPB	1/12/2024 15:48	Std
	10 S05	46211	-		10 SFAM01.1	PPB	1/12/2024 15:55	Std
ICV065	ICV065	19886	4.2354 -		SFAM01.1	PPB	1/12/2024 15:57	SMPL
ICB065	ICB065	-118	-0.0636 -		SFAM01.1	PPB	1/12/2024 16:00	SMPL
CCV069	CCV069	23648	5.0439 -		SFAM01.1	PPB	1/12/2024 16:05	SMPL
CCB069	CCB069	-146	-0.0696 -		SFAM01.1	PPB	1/12/2024 16:07	SMPL
PB158422BL	PBW422	-38	-0.0464 -		SFAM01.1	PPB	1/12/2024 16:09	SMPL
PB158367TB	LEB367	-19	-0.0423 -		SFAM01.1	PPB	1/12/2024 16:11	SMPL
P1044-01	EOX92	9	-0.0363 -		SFAM01.1	PPB	1/12/2024 16:14	SMPL
P1044-02	EOX78	438	0.0559 -		SFAM01.1	PPB	1/12/2024 16:16	SMPL
P1044-03	EOX78D	402	0.0482 -		SFAM01.1	PPB	1/12/2024 16:23	SMPL
P1044-04	EOX78S	6864	1.4369 -		SFAM01.1	PPB	1/12/2024 16:26	SMPL
CCV070	CCV070	23653	5.045 -		SFAM01.1	PPB	1/12/2024 16:28	SMPL
CCB070	CCB070	-34	-0.0455 -		SFAM01.1	PPB	1/12/2024 16:30	SMPL