



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129144

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/15/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	6.98	01/15/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	01/15/2024	08:19
4	ICV	1	Water	NA	NA	20.4	7.00	01/15/2024	08:22
5	CCV1	1	Water	NA	NA	20.3	2.01	01/15/2024	08:25
6	P1133-01	1	Water	NA	NA	24.4	5.89	01/15/2024	08:35
7	P1133-01DUP	1	Water	NA	NA	24.5	5.90	01/15/2024	08:36
8	CCV2	1	Water	NA	NA	20.3	12.02	01/15/2024	08:40

129146

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1133

WorkList ID : 177160

Department : Wet-Chemistry

Date : 01-15-2024 08:01:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1133-01	TOTE-6579	Water	pH	Cool 4 deg C	PSEG03	O41	01/12/2024	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01-15-24 08:10
JD (029)
JD (CSM)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01-15-24 10:00
JD (CSM)
JD (00C)