

LB129145


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43514



A= 0.0000e+000

B= 2.2560e-004

C= 2.0971e-002

Rho= 0.9990540

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/11)
0.0	0.000	0.035	0.035	62	0.000	62	0				
0.2	0.200	0.186	-0.014	733	0.0 %	733					7
2.5	2.500	2.283	-0.217	10028	0.0 %	10028					9
5.0	5.000	5.268	0.268	23258	0.0 %	23258					5
7.5	7.500	7.589	0.089	33548	0.0 %	33548					1
10.0	10.000	9.838	-0.162	43514	0.0 %	43514					2

LB129145 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Jan 2024 09:54:17

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	62	-	0.0000	7470A	PPB	15 Jan 2024 10:14:24	S	Std
0.2 - 1	S0-2	733	-	0.2000	7470A	PPB	15 Jan 2024 10:16:48	S	Std
2.5 - 1	S2-5	10028	-	2.5000	7470A	PPB	15 Jan 2024 10:19:06	S	Std
5.0 - 1	S5	23258	-	5.0000	7470A	PPB	15 Jan 2024 10:21:22	S	Std
7.5 - 1	S7-5	33548	-	7.5000	7470A	PPB	15 Jan 2024 10:23:41	S	Std
10.0 - 1	S10	43514	-	10.0000	7470A	PPB	15 Jan 2024 10:25:59	S	Std
ICV66 - 1	ICV66	16876	3.8282	-	7470A	PPB	15 Jan 2024 10:31:29	U	SMPL
ICB66 - 1	ICB66	-234	-0.0318	-	7470A	PPB	15 Jan 2024 10:33:44	U	SMPL
CCV71 - 1	CCV71	22488	5.0943	-	7470A	PPB	15 Jan 2024 10:36:01	U	SMPL
CCB71 - 1	CCB71	-5	0.0198	-	7470A	PPB	15 Jan 2024 10:43:39	U	SMPL
CRA - 1	CRA	700	0.1789	-	7470A	PPB	15 Jan 2024 10:45:54	U	SMPL
HighStd - 1	HighStd	42725	9.6599	-	7470A	PPB	15 Jan 2024 10:48:10	U	SMPL
ChkStd - 1	ChkStd	31919	7.2220	-	7470A	PPB	15 Jan 2024 10:50:25	U	SMPL
PB158415BL - 1	PBW	-389	-0.0668	-	7470A	PPB	15 Jan 2024 10:52:45	U	SMPL
PB158415BS - 1	LCSW	15503	3.5185	-	7470A	PPB	15 Jan 2024 10:55:03	U	SMPL
P1104-04 - 1	TP-1	-211	-0.0266	-	7470A	PPB	15 Jan 2024 10:57:19	U	SMPL
P1104-08 - 1	TP-2	-18	0.0169	-	7470A	PPB	15 Jan 2024 10:59:39	U	SMPL
P1117-02 - 1	S1-COMPOSITE	-12	0.0183	-	7470A	PPB	15 Jan 2024 11:01:58	U	SMPL
P1117-08 - 1	S2-COMPOSITE	75	0.0379	-	7470A	PPB	15 Jan 2024 11:04:15	U	SMPL
P1118-04 - 1	WC-1	127	0.0496	-	7470A	PPB	15 Jan 2024 11:06:30	U	SMPL
CCV72 - 1	CCV72	22762	5.1561	-	7470A	PPB	15 Jan 2024 11:08:45	U	SMPL
CCB72 - 1	CCB72	-288	-0.0440	-	7470A	PPB	15 Jan 2024 11:11:02	U	SMPL
P1118-08 - 1	WC-2	69	0.0365	-	7470A	PPB	15 Jan 2024 11:13:20	U	SMPL
P1118-08DUP - 1	WC-2DUP	62	0.0350	-	7470A	PPB	15 Jan 2024 11:15:35	U	SMPL
P1118-08MS - 1	WC-2MS	17014	3.8594	-	7470A	PPB	15 Jan 2024 11:17:51	U	SMPL
P1118-08MSD - 1	WC-2MSD	17790	4.0344	-	7470A	PPB	15 Jan 2024 11:20:06	U	SMPL
PB158416BL - 1	PBW	-258	-0.0372	-	7470A	PPB	15 Jan 2024 11:22:24	U	SMPL
PB158416BS - 1	LCSW	16956	3.8463	-	7470A	PPB	15 Jan 2024 11:24:43	U	SMPL
P1098-04 - 1	AUD-1311	-174	-0.0183	-	7470A	PPB	15 Jan 2024 11:26:59	U	SMPL
P1098-04DUP - 1	AUD-1311DUP	25	0.0266	-	7470A	PPB	15 Jan 2024 11:29:19	U	SMPL
P1098-04MS - 1	AUD-1311MS	17867	4.0518	-	7470A	PPB	15 Jan 2024 11:31:35	U	SMPL
P1098-04MSD - 1	AUD-1311MSD	18494	4.1933	-	7470A	PPB	15 Jan 2024 11:33:51	U	SMPL
CCV73 - 1	CCV73	23273	5.2714	-	7470A	PPB	15 Jan 2024 11:36:09	U	SMPL
CCB73 - 1	CCB73	-300	-0.0467	-	7470A	PPB	15 Jan 2024 11:38:27	U	SMPL
PB158417BL - 1	PBW	-56	0.0083	-	7470A	PPB	15 Jan 2024 11:40:45	U	SMPL
PB158417BS - 1	LCSW	16570	3.7592	-	7470A	PPB	15 Jan 2024 11:43:00	U	SMPL
P1106-01 - 1	34495	341	0.0979	-	7470A	PPB	15 Jan 2024 11:50:03	U	SMPL
P1106-01DUP - 1	34495DUP	-83	0.0022	-	7470A	PPB	15 Jan 2024 11:52:19	U	SMPL
P1106-01MS - 1	34495MS	17251	3.9128	-	7470A	PPB	15 Jan 2024 11:54:34	U	SMPL
P1106-01MSD - 1	34495MSD	16699	3.7883	-	7470A	PPB	15 Jan 2024 11:59:32	U	SMPL
P1133-01 - 1	TOTE-6579	883	0.2202	-	7470A	PPB	15 Jan 2024 12:01:48	U	SMPL
PB158388TB - 1	PB158388TB	-81	0.0027	-	7470A	PPB	15 Jan 2024 12:04:07	U	SMPL
PB158394TB - 1	PB158394TB	27	0.0271	-	7470A	PPB	15 Jan 2024 12:06:24	U	SMPL
PB158169TB - 1	PB158169TB	41	0.0302	-	7470A	PPB	15 Jan 2024 12:08:40	U	SMPL
CCV74 - 1	CCV74	21934	4.9693	-	7470A	PPB	15 Jan 2024 12:10:55	U	SMPL
CCB74 - 1	CCB74	-143	-0.0113	-	7470A	PPB	15 Jan 2024 12:13:11	U	SMPL
P1118-08LX5 - 1		5	0.0221	-	7470A	PPB	15 Jan 2024 12:15:28	U	SMPL
P1118-08A - 1		17449	3.9575	-	7470A	PPB	15 Jan 2024 12:17:44	U	SMPL
P1098-04LX5 - 1		-152	-0.0133	-	7470A	PPB	15 Jan 2024 12:19:59	U	SMPL
P1098-04A - 1		17750	4.0254	-	7470A	PPB	15 Jan 2024 12:22:18	U	SMPL
P1106-01LX5 - 1		-58	0.0079	-	7470A	PPB	15 Jan 2024 12:27:32	U	SMPL
P1106-01A - 1		16980	3.8517	-	7470A	PPB	15 Jan 2024 12:29:48	U	SMPL
CCV75 - 1	CCV75	22763	5.1564	-	7470A	PPB	15 Jan 2024 12:32:05	U	SMPL
CCB75 - 1	CCB75	-183	-0.0203	-	7470A	PPB	15 Jan 2024 12:34:24	U	SMPL