



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129160

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/16/2024	15:40
2	CAL2	1	Water	NA	NA	20.3	7.00	01/16/2024	15:41
3	CAL3	1	Water	NA	NA	20.4	10.02	01/16/2024	15:44
4	ICV	1	Water	NA	NA	20.3	7.01	01/16/2024	15:47
5	CCV1	1	Water	NA	NA	20.3	2.01	01/16/2024	15:50
6	P1151-01	1	Water	NA	NA	20.8	6.36	01/16/2024	16:00
7	P1152-01	1	Water	NA	NA	20.7	6.73	01/17/2024	16:05
8	P1152-01DUP	1	Water	NA	NA	20.8	6.74	01/16/2024	16:06
9	CCV2	1	Water	NA	NA	20.5	12.02	01/16/2024	16:11

UP 129160

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1152 WorkList ID : 177196 Department : Wet-Chemistry Date : 01-16-2024 15:22:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1151-01	JS10000	Water	pH	Cool 4 deg C	STER14	D11	01/12/2024	9040C
P1152-01	A2421	Water	pH	Cool 4 deg C	PSEG03	E21	01/16/2024	9040C

Date/Time 01-16-24 15:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 01-16-24 18:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]