


LB129161

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9019e-004 slope

C= -1.3030e-001 *intercept*

Rho= 0.9999659

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/I)
0.00	0.000	0.044	0.044	919	0.000	919					
0.05	0.050					442					
0.20	0.200	0.168	-0.032	1568	0.0 %	1568	0				-16
2.50	2.500	2.484	-0.016	13747	0.0 %	13747					-1
5.00	5.000	5.018	0.018	27068	0.0 %	27068					0
7.50	7.500	7.462	-0.038	39920	0.0 %	39920					-1
10.0	10.000	10.024	0.024	53389	0.0 %	53389					0

LB129161

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	919	-		0 SFAM01.1	PPB	1/17/2024 10:47	Std
	0.2 S01	1568	-		0.2 SFAM01.1	PPB	1/17/2024 10:49	Std
	2.5 S02	13747	-		2.5 SFAM01.1	PPB	1/17/2024 10:51	Std
	5 S03	27068	-		5 SFAM01.1	PPB	1/17/2024 10:54	Std
	7.5 S04	39920	-		7.5 SFAM01.1	PPB	1/17/2024 10:56	Std
	10 S05	53389	-		10 SFAM01.1	PPB	1/17/2024 11:01	Std
ICV068	ICV068	20705	3.8075 -		SFAM01.1	PPB	1/17/2024 11:04	SMPL
ICB068	ICB068	722	0.007 -		SFAM01.1	PPB	1/17/2024 11:06	SMPL
CCV079	CCV079	26085	4.8308 -		SFAM01.1	PPB	1/17/2024 11:09	SMPL
CCB079	CCB079	705	0.0038 -		SFAM01.1	PPB	1/17/2024 11:11	SMPL
PB158467BL	PBS467	656	-0.0055 -		SFAM01.1	PPB	1/17/2024 11:13	SMPL
P1022-02	E0X95	2477	0.3408 -		SFAM01.1	PPB	1/17/2024 11:15	SMPL
P1022-03	E0X96	1175	0.0932 -		SFAM01.1	PPB	1/17/2024 11:18	SMPL
P1022-04	E0X97	1097	0.0783 -		SFAM01.1	PPB	1/17/2024 11:20	SMPL
P1022-05	E0X98	499	-0.0354 -		SFAM01.1	PPB	1/17/2024 11:22	SMPL
P1022-06	E0XN1	527	-0.0301 -		SFAM01.1	PPB	1/17/2024 11:25	SMPL
P1022-07	E0X91	804	0.0226 -		SFAM01.1	PPB	1/17/2024 11:27	SMPL
P1022-08	E0X92	314	-0.0706 -		SFAM01.1	PPB	1/17/2024 11:29	SMPL
P1022-09	E0X93	644	-0.0078 -		SFAM01.1	PPB	1/17/2024 11:31	SMPL
P1022-10	E0X93D	716	0.0059 -		SFAM01.1	PPB	1/17/2024 11:34	SMPL
P1022-11	E0X93S	16993	3.1016 -		SFAM01.1	PPB	1/17/2024 11:36	SMPL
P1022-12	E0X79	303	-0.0727 -		SFAM01.1	PPB	1/17/2024 11:38	SMPL
P1022-13	E0X80	281	-0.0769 -		SFAM01.1	PPB	1/17/2024 11:40	SMPL
P1022-14	E0X81	14500	2.6274 -		SFAM01.1	PPB	1/17/2024 11:43	SMPL
P1022-15	E0X82	977	0.0555 -		SFAM01.1	PPB	1/17/2024 11:45	SMPL
P1022-16	E0XN2	317	-0.07 -		SFAM01.1	PPB	1/17/2024 11:47	SMPL
P1022-20	E0XH6	4721	0.7676 -		SFAM01.1	PPB	1/17/2024 11:50	SMPL
P1022-21	E0XH7	255	-0.0818 -		SFAM01.1	PPB	1/17/2024 11:52	SMPL
P1022-22	E0XX8	410	-0.0523 -		SFAM01.1	PPB	1/17/2024 11:54	SMPL
P1022-23	E0X75	6772	1.1577 -		SFAM01.1	PPB	1/17/2024 11:56	SMPL
CCV080	CCV080	25496	4.7187 -		SFAM01.1	PPB	1/17/2024 11:59	SMPL
CCB080	CCB080	744	0.0112 -		SFAM01.1	PPB	1/17/2024 12:01	SMPL
P1022-24	E0X76	3681	0.5698 -		SFAM01.1	PPB	1/17/2024 12:03	SMPL
P1022-25	E0X77	416	-0.0512 -		SFAM01.1	PPB	1/17/2024 12:05	SMPL
PB158468BL	PBS468	653	-0.0061 -		SFAM01.1	PPB	1/17/2024 12:08	SMPL
P1065-01	E0X94	519	-0.0316 -		SFAM01.1	PPB	1/17/2024 12:10	SMPL
P1065-02	E0XZ4	766	0.0154 -		SFAM01.1	PPB	1/17/2024 12:12	SMPL
P1065-03	E0XZ7	207	-0.0909 -		SFAM01.1	PPB	1/17/2024 12:15	SMPL
P1065-04	E0XH5	9332	1.6445 -		SFAM01.1	PPB	1/17/2024 12:17	SMPL
P1065-05	E0XH5D	7952	1.3821 -		SFAM01.1	PPB	1/17/2024 12:19	SMPL
P1065-06	E0XH5S	20426	3.7545 -		SFAM01.1	PPB	1/17/2024 12:21	SMPL
P1065-08	E0X67	12168	2.1839 -		SFAM01.1	PPB	1/17/2024 12:24	SMPL
P1065-09	E0X78	71	-0.1168 -		SFAM01.1	PPB	1/17/2024 12:26	SMPL
P1065-10	E0Y05	1733	0.1993 -		SFAM01.1	PPB	1/17/2024 12:28	SMPL
P1065-11	E0X68	229	-0.0867 -		SFAM01.1	PPB	1/17/2024 12:31	SMPL
P1065-12	E0X69	207	-0.0909 -		SFAM01.1	PPB	1/17/2024 12:33	SMPL

LB129161

INSTRUMENT ID : CV1

P1065-13	EOX70	831	0.0277 -	SFAM01.1	PPB	1/17/2024 12:35 SMPL
P1065-14	EOY06	1140	0.0865 -	SFAM01.1	PPB	1/17/2024 12:37 SMPL
P1065-15	EOY07	548	-0.0261 -	SFAM01.1	PPB	1/17/2024 12:40 SMPL
P1065-16	EOX63	838	0.0291 -	SFAM01.1	PPB	1/17/2024 12:42 SMPL
P1065-17	EOY11	795	0.0209 -	SFAM01.1	PPB	1/17/2024 12:44 SMPL
CCV081	CCV081	25258	4.6735 -	SFAM01.1	PPB	1/17/2024 12:46 SMPL
CCB081	CCB081	730	0.0085 -	SFAM01.1	PPB	1/17/2024 12:49 SMPL