

LB129165


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

57895



A= 0.0000e+000

B= 1.7323e-004

C= -1.9417e-001

Rho= 0.9984806

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.010	-0.010	1066	0.000	1066					
0.2	0.200	0.173	-0.027	2118	0.0 %	2118	0				-14
2.5	2.500	2.667	0.167	16517	0.0 %	16517					7
5.0	5.000	4.708	-0.292	28300	0.0 %	28300					-6
7.5	7.500	7.827	0.327	46304	0.0 %	46304					4
10.0	10.000	9.835	-0.165	57895	0.0 %	57895					-2

LB129165 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jan 2024 15:15:29

Sample ID	Extended ID	u Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1066	-	0.0000	7470A	PPB	17 Jan 2024 15:46:43	S	Std
0.2 - 1	50-2	2118	-	0.2000	7470A	PPB	17 Jan 2024 15:51:13	S	Std
2.5 - 1	52-1	16517	-	2.5000	7470A	PPB	17 Jan 2024 15:53:30	S	Std
5.0 - 1	55	28300	-	5.0000	7470A	PPB	17 Jan 2024 15:55:47	S	Std
7.5 - 1	57-1	46304	-	7.5000	7470A	PPB	17 Jan 2024 16:00:50	S	Std
10.0 - 1	510	57895	-	10.0000	7470A	PPB	17 Jan 2024 16:08:29	S	Std
ICV69 - 1	ICV69	24440	4.0395	-	7470A	PPB	17 Jan 2024 16:11:24	U	SMPL
ICB69 - 1	ICB69	1277	0.0270	-	7470A	PPB	17 Jan 2024 16:16:16	U	SMPL
CCV82 - 1	CCV82	28430	4.7306	-	7470A	PPB	17 Jan 2024 16:18:31	U	SMPL
CCB82 - 1	CCB82	1224	0.0179	-	7470A	PPB	17 Jan 2024 16:20:47	U	SMPL
CRA - 1	CRA	2118	0.1727	-	7470A	PPB	17 Jan 2024 16:23:04	U	SMPL
HighStd - 1	HighStd	58333	9.9106	-	7470A	PPB	17 Jan 2024 16:25:19	U	SMPL
ChkStd - 1	ChkStd	44003	7.4283	-	7470A	PPB	17 Jan 2024 16:27:34	U	SMPL
PB158465BL - 1	PBW	960	-0.0279	-	7470A	PPB	17 Jan 2024 16:32:35	U	SMPL
PB158465BS - 1	LCSW	27149	4.5087	-	7470A	PPB	17 Jan 2024 16:34:51	U	SMPL
P1139-02 - 1	MR-314-0-2	877	-0.0422	-	7470A	PPB	17 Jan 2024 16:41:59	U	SMPL
P1139-04 - 1	MR-314-16-18	1030	-0.0157	-	7470A	PPB	17 Jan 2024 16:44:15	U	SMPL
P1140-02 - 1	MR-309-0-2	1257	0.0236	-	7470A	PPB	17 Jan 2024 16:46:32	U	SMPL
P1140-04 - 1	MR-309-37-39	347	-0.1341	-	7470A	PPB	17 Jan 2024 16:48:48	U	SMPL
P1140-04DUP - 1	MR-309-37-39DUP	939	-0.0315	-	7470A	PPB	17 Jan 2024 16:53:39	U	SMPL
CCV83 - 1	CCV83	28644	4.7677	-	7470A	PPB	17 Jan 2024 16:55:54	U	SMPL
CCB83 - 1	CCB83	1244	0.0213	-	7470A	PPB	17 Jan 2024 16:58:09	U	SMPL
P1140-04MS - 1	MR-309-37-39MS	28402	4.7258	-	7470A	PPB	17 Jan 2024 17:00:26	U	SMPL
P1140-04MSD - 1	MR-309-37-39MSD	27902	4.6392	-	7470A	PPB	17 Jan 2024 17:05:21	U	SMPL
P1140-04LX5 - 1		46	-0.1862	-	7470A	PPB	17 Jan 2024 17:09:54	U	SMPL
P1140-04A - 1		24305	4.0161	-	7470A	PPB	17 Jan 2024 17:12:10	U	SMPL
CCV84 - 1	CCV84	28617	4.7630	-	7470A	PPB	17 Jan 2024 17:14:26	U	SMPL
CCB84 - 1	CCB84	625	-0.0859	-	7470A	PPB	17 Jan 2024 17:16:45	U	SMPL