



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129167

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	01/18/2024	11:30
2	P1152-01		1	13	100.00	>212.0	>212.0	01/18/2024	12:00
3	P1152-01DUP		1	14	100.00	>212.0	>212.0	01/18/2024	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB129167

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint 01/18/2024 WorkList ID : 177240 Department : Wet-Chemistry Date : 01-18-2024 10:56:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1152-01	A2421	Water	Flash Point	Cool 4 deg C	PSEG03	E21	01/16/2024	1010B

Date/Time 01/18/2024 11:00
Raw Sample Received by: NP W-C
Raw Sample Relinquished by: NP W-C

Date/Time 01/18/2024 13:10
Raw Sample Received by: NP W-C
Raw Sample Relinquished by: NP W-C